


LB119837

7470A

INSTRUMENT ID: CVZ



Linear

A= 0.0000e+000

B= 2.3360e-004

C= 4.2726e-002

Rho= 0.9972126

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.021	0.021	-91	0.000	-91					-
0.2	0.200	0.223	0.023	772	0.0 %	772					12
2.5	2.500	2.164	-0.336	9080	0.0 %	9080					-13
5.0	5.000	5.366	0.366	22788	0.0 %	22788					7
7.5	7.500	7.782	0.282	33132	0.0 %	33132					4
10.0	10.000	9.643	-0.357	41099	0.0 %	41099					-4

LB119837 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Apr 2022 09:36:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-91	-	0.0000	7470A	PPB	26 Apr 2022 10:42:10	S	Std
0.2 - 1	50.2	772	-	0.2000	7470A	PPB	26 Apr 2022 10:47:16	S	Std
2.5 - 1	52.5	9080	-	2.5000	7470A	PPB	26 Apr 2022 10:49:33	S	Std
5.0 - 1	55	22788	-	5.0000	7470A	PPB	26 Apr 2022 10:51:49	S	Std
7.5 - 1	57.5	33132	-	7.5000	7470A	PPB	26 Apr 2022 10:54:05	S	Std
10.0 - 1	510	41099	-	10.0000	7470A	PPB	26 Apr 2022 10:56:21	S	Std
ICV93 - 1	ICV93	16059	3.7941	-	7470A	PPB	26 Apr 2022 10:59:10	U	SMPL
ICB93 - 1	ICB93	-97	0.0201	-	7470A	PPB	26 Apr 2022 11:01:24	U	SMPL
CCV29 - 1	CCV29	21515	5.0686	-	7470A	PPB	26 Apr 2022 11:03:39	U	SMPL
CCB29 - 1	CCB29	-82	0.0236	-	7470A	PPB	26 Apr 2022 11:05:54	U	SMPL
CRA - 1	CRA	778	0.2245	-	7470A	PPB	26 Apr 2022 11:08:09	U	SMPL
HighStd - 1	HighStd	41382	9.7095	-	7470A	PPB	26 Apr 2022 11:10:24	U	SMPL
ChkStd - 1	ChkStd	31169	7.3238	-	7470A	PPB	26 Apr 2022 11:12:39	U	SMPL
PB144350BL - 1	PBW	-117	0.0154	-	7470A	PPB	26 Apr 2022 11:14:55	U	SMPL
PB144350BS - 1	LCSW	17682	4.1732	-	7470A	PPB	26 Apr 2022 11:17:12	U	SMPL
N2548-02 - 1	MH-8S	-72	0.0259	-	7470A	PPB	26 Apr 2022 11:19:28	U	SMPL
N2548-05 - 1	MH-8D	7	0.0444	-	7470A	PPB	26 Apr 2022 11:21:45	U	SMPL
N2548-08 - 1	TP-9	19	0.0472	-	7470A	PPB	26 Apr 2022 11:24:02	U	SMPL
N2548-11 - 1	TP-11	15	0.0462	-	7470A	PPB	26 Apr 2022 11:26:19	U	SMPL
N2548-11DUP - 1	TP-11DUP	-2	0.0423	-	7470A	PPB	26 Apr 2022 11:28:34	U	SMPL
CCV30 - 1	CCV30	21238	5.0039	-	7470A	PPB	26 Apr 2022 11:30:50	U	SMPL
CCB30 - 1	CCB30	-85	0.0229	-	7470A	PPB	26 Apr 2022 11:33:05	U	SMPL
N2548-11MS - 1	TP-11MS	18238	4.3031	-	7470A	PPB	26 Apr 2022 11:35:20	U	SMPL
N2548-11MSD - 1	TP-11MSD	18260	4.3082	-	7470A	PPB	26 Apr 2022 11:37:35	U	SMPL
PB144329TB - 1	PB144329TB	-90	0.0217	-	7470A	PPB	26 Apr 2022 11:39:50	U	SMPL
N2548-11LX5 - 1		-14	0.0395	-	7470A	PPB	26 Apr 2022 11:42:06	U	SMPL
N2548-11A - 1		17045	4.0244	-	7470A	PPB	26 Apr 2022 11:44:22	U	SMPL
CCV31 - 1	CCV31	21500	5.0651	-	7470A	PPB	26 Apr 2022 11:46:38	U	SMPL
CCB31 - 1	CCB31	-105	0.0182	-	7470A	PPB	26 Apr 2022 11:48:54	U	SMPL