


7471B

LB119838

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 1.8155e-004

C= 6.5985e-004

Rho= 0.9981716

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	125	0.000	125					---
0.2	0.200	0.244	0.044	1343	0.0 %	1343					22
2.5	2.500	2.276	-0.224	12531	0.0 %	12531					-9
5.0	5.000	5.068	0.068	27911	0.0 %	27911					1
7.5	7.500	7.889	0.389	43447	0.0 %	43447					5
10.0	10.000	9.700	-0.300	53424	0.0 %	53424					-3

LB119838 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Apr 2022 12:06:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	125	-	0.0000	7471B	PPB	26 Apr 2022 12:21:14	S	Std
0.2 - 1	50.2	1343	-	0.2000	7471B	PPB	26 Apr 2022 12:25:28	S	Std
2.5 - 1	52.5	12531	-	2.5000	7471B	PPB	26 Apr 2022 12:32:33	S	Std
5.0 - 1	55	27911	-	5.0000	7471B	PPB	26 Apr 2022 12:34:50	S	Std
7.5 - 1	57.5	43447	-	7.5000	7471B	PPB	26 Apr 2022 12:37:07	S	Std
10.0 - 1	510	53424	-	10.0000	7471B	PPB	26 Apr 2022 12:47:25	S	Std
ICV94 - 1	ICV94	21876	3.9723	-	7471B	PPB	26 Apr 2022 12:50:25	U	SMPL
ICB94 - 1	ICB94	182	0.0337	-	7471B	PPB	26 Apr 2022 12:52:39	U	SMPL
CCV32 - 1	CCV32	27517	4.9964	-	7471B	PPB	26 Apr 2022 12:54:55	U	SMPL
CCB32 - 1	CCB32	181	0.0335	-	7471B	PPB	26 Apr 2022 12:57:10	U	SMPL
CRA - 1	CRA	1303	0.2372	-	7471B	PPB	26 Apr 2022 12:59:25	U	SMPL
HighStd - 1	HighStd	55882	10.1462	-	7471B	PPB	26 Apr 2022 13:01:41	U	SMPL
ChkStd - 1	ChkStd	39752	7.2177	-	7471B	PPB	26 Apr 2022 13:03:56	U	SMPL
PB144374BL - 1	PBS	-71	-0.0122	-	7471B	PPB	26 Apr 2022 13:06:14	U	SMPL
PB144374BS - 1	LCSS	20643	3.7484	-	7471B	PPB	26 Apr 2022 13:08:31	U	SMPL
N2502-01 - 1	C1	706	0.1288	-	7471B	PPB	26 Apr 2022 13:10:48	U	SMPL
N2502-03 - 1	C2	1625	0.2957	-	7471B	PPB	26 Apr 2022 13:13:04	U	SMPL
N2502-05 - 1	C3	1750	0.3184	-	7471B	PPB	26 Apr 2022 13:15:20	U	SMPL
N2502-07 - 1	C4	1248	0.2272	-	7471B	PPB	26 Apr 2022 13:17:37	U	SMPL
N2502-09 - 1	C5	905	0.1650	-	7471B	PPB	26 Apr 2022 13:19:52	U	SMPL
CCV33 - 1	CCV33	27828	5.0529	-	7471B	PPB	26 Apr 2022 13:22:07	U	SMPL
CCB33 - 1	CCB33	175	0.0324	-	7471B	PPB	26 Apr 2022 13:24:23	U	SMPL
N2541-01 - 1	CL-02-042222	2681	0.4874	-	7471B	PPB	26 Apr 2022 13:26:39	U	SMPL
N2548-01 - 1	MH-8S	640	0.1169	-	7471B	PPB	26 Apr 2022 13:28:53	U	SMPL
N2548-01DUP - 1	MH-8SDUP	655	0.1196	-	7471B	PPB	26 Apr 2022 13:31:09	U	SMPL
N2548-01MS - 1	MH-8SMS	21218	3.8528	-	7471B	PPB	26 Apr 2022 13:33:24	U	SMPL
N2548-01MSD - 1	MH-8SMSD	21653	3.9318	-	7471B	PPB	26 Apr 2022 13:35:40	U	SMPL
N2548-04 - 1	MH-8D	622	0.1136	-	7471B	PPB	26 Apr 2022 13:37:56	U	SMPL
N2548-07 - 1	TP-9	1354	0.2465	-	7471B	PPB	26 Apr 2022 13:40:13	U	SMPL
N2548-10 - 1	TP-11	12511	2.2721	-	7471B	PPB	26 Apr 2022 13:42:30	U	SMPL
N2553-01 - 1	370	1728	0.3144	-	7471B	PPB	26 Apr 2022 13:44:46	U	SMPL
N2554-01 - 1	RBR-2520253	965	0.1759	-	7471B	PPB	26 Apr 2022 13:47:02	U	SMPL
CCV34 - 1	CCV34	27383	4.9721	-	7471B	PPB	26 Apr 2022 13:49:17	U	SMPL
CCB34 - 1	CCB34	203	0.0375	-	7471B	PPB	26 Apr 2022 13:51:32	U	SMPL
N2555-01 - 1	OR-03-042522	3516	0.6390	-	7471B	PPB	26 Apr 2022 13:53:48	U	SMPL
N2559-01 - 1	MH-10S	1033	0.1882	-	7471B	PPB	26 Apr 2022 13:56:03	U	SMPL
N2559-04 - 1	MH-10D	1166	0.2124	-	7471B	PPB	26 Apr 2022 14:01:35	U	SMPL
N2548-01LX5 - 1		264	0.0486	-	7471B	PPB	26 Apr 2022 14:03:51	U	SMPL
N2548-01A - 1		21704	3.9411	-	7471B	PPB	26 Apr 2022 14:06:06	U	SMPL
CCV35 - 1	CCV35	27787	5.0455	-	7471B	PPB	26 Apr 2022 14:08:22	U	SMPL
CCB35 - 1	CCB35	213	0.0393	-	7471B	PPB	26 Apr 2022 14:10:39	U	SMPL