



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB119839

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/26/2022	10:05
2	N2550-01		1	12	100.00	>212.0	>212.0	04/26/2022	10:25
3	N2550-01DUP		1	12	100.00	>212.0	>212.0	04/26/2022	10:40
4	N2550-02		1	15	100.00	>212.0	>212.0	04/26/2022	11:10
5	N2552-03		1	15	100.00	>212.0	>212.0	04/26/2022	11:35
6	N2552-05		1	14	100.00	>212.0	>212.0	04/26/2022	11:55
7	N2557-02		1	12	100.00	>212.0	>212.0	04/26/2022	12:20
8	N2558-01		1	14	100.00	>212.0	>212.0	04/26/2022	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB119839

WorkList Name : fp42622

WorkList ID : 159009

Department : Wet-Chemistry

Date : 04-26-2022 09:36:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/02/2022	Water	N2550-01	Flash Point	Cool 4 deg C	DIST01	P11	PUMP-STATION-207ST-TRAIN-FACILITY	04/23/2022	1010B
05/02/2022	Water	N2550-02	Flash Point	Cool 4 deg C	DIST01	P11	DYCKMAN-200ST-A51515T50SB	04/23/2022	1010B
05/02/2022	Water	N2552-03	Flash Point	Cool 4 deg C	PSEG03	K11	33640	04/25/2022	1010B
05/02/2022	Water	N2552-05	Flash Point	Cool 4 deg C	PSEG03	K11	33645	04/25/2022	1010B
05/09/2022	Water	N2557-02	Flash Point	Cool 4 deg C	SEAC01	L11	PURGE-WATER	04/25/2022	1010B
05/02/2022	Water	N2558-01	Flash Point	Cool 4 deg C	PSEG03	K11	31699	04/25/2022	1010B

Date/Time 04/26/22 09:45
 Raw Sample Received by: KS (wq) (900)
 Raw Sample Relinquished by: KS (wq) (900)

Date/Time 04/26/22 12:45
 Raw Sample Received by: KS (wq) (900)
 Raw Sample Relinquished by: KS (wq) (900)