



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119843

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	04/26/2022	13:10
2	CAL2	1	Water	NA	NA	20.8	7.00	04/26/2022	13:11
3	CAL3	1	Water	NA	NA	20.7	10.03	04/26/2022	13:12
4	ICV	1	Water	NA	NA	20.7	6.99	04/26/2022	13:15
5	CCV1	1	Water	NA	NA	20.7	2.01	04/26/2022	13:16
6	N2563-01	1	Water	NA	NA	20.2	7.84	04/26/2022	13:19
7	N2563-01DUP	1	Water	NA	NA	20.3	7.85	04/26/2022	13:20
8	CCV2	1	Water	NA	NA	20.6	12.03	04/26/2022	13:22

VB 119843

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2563 WorkList ID : 159014 Department : Wet-Chemistry Date : 04-26-2022 12:51:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/26/2022	Water	N2563-01	pH	Cool 4 deg C	ENTA05	K11	STORM-WATER-042622	04/26/2022	SM 4500-H B

Date/Time 04/26/22 23:00
Raw Sample Received by: TP(WC)
Raw Sample Relinquished by: CP SM

Date/Time 04/26/22 14:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: TP(WC)