

LB119848


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43562



A= 0.0000e+000

B= 2.2051e-004 *slope*C= 9.1997e-004 *y-intercept*

Rho= 0.9971800

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%DI
0.00	0.000	0.003	0.003	10	0.000	10					=
0.05	0.050					185	0				22
0.20	0.200	0.243	0.043	1099	0.0 %	1099					-12
2.50	2.500	2.206	-0.294	10002	0.0 %	10002					5
5.00	5.000	5.242	0.242	23769	0.0 %	23769					5
7.50	7.500	7.898	0.398	35815	0.0 %	35815					-4
10.0	10.000	9.607	-0.393	43562	0.0 %	43562					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	10	-		0 SFAM01.1	PPB	4/26/2022 15:45	Std
	0.2 S01	1099	-		0.2 SFAM01.1	PPB	4/26/2022 15:47	Std
	2.5 S02	10002	-		2.5 SFAM01.1	PPB	4/26/2022 15:54	Std
	5 S03	23769	-		5 SFAM01.1	PPB	4/26/2022 15:57	Std
	7.5 S04	35815	-		7.5 SFAM01.1	PPB	4/26/2022 15:59	Std
	10 S05	43562	-		10 SFAM01.1	PPB	4/26/2022 16:01	Std
ICV095	ICV095	17819	3.9301 -		SFAM01.1	PPB	4/26/2022 16:06	SMPL
ICB095	ICB095	-48	-0.0097 -		SFAM01.1	PPB	4/26/2022 16:08	SMPL
CCV036	CCV036	23774	5.2433 -		SFAM01.1	PPB	4/26/2022 16:10	SMPL
CCB036	CCB036	-73	-0.0152 -		SFAM01.1	PPB	4/26/2022 16:13	SMPL
PB144377BL	PBW377	-3	0.0003 -		SFAM01.1	PPB	4/26/2022 16:15	SMPL
N2264-01	YAZ17	4	0.0018 -		SFAM01.1	PPB	4/26/2022 16:17	SMPL
N2264-02	YAZ17D	1	0.0011 -		SFAM01.1	PPB	4/26/2022 16:19	SMPL
N2264-03	YAZ17S	3865	0.8532 -		SFAM01.1	PPB	4/26/2022 16:22	SMPL
N2264-04	YAZ15	273	0.0611 -		SFAM01.1	PPB	4/26/2022 16:24	SMPL
N2264-05	YAZ21	58	0.0137 -		SFAM01.1	PPB	4/26/2022 16:26	SMPL
N2264-06	YAZ13	45	0.0108 -		SFAM01.1	PPB	4/26/2022 16:28	SMPL
N2264-07	YAYW1	40	0.0097 -		SFAM01.1	PPB	4/26/2022 16:31	SMPL
N2264-08	YAZ23	16	0.0044 -		SFAM01.1	PPB	4/26/2022 16:33	SMPL
N2264-09	YAZ67	7	0.0025 -		SFAM01.1	PPB	4/26/2022 16:35	SMPL
N2264-10	YAYW3	325	0.0726 -		SFAM01.1	PPB	4/26/2022 16:37	SMPL
N2264-11	YAYW5	344	0.0768 -		SFAM01.1	PPB	4/26/2022 16:40	SMPL
N2264-12	YAYZ9	447	0.0995 -		SFAM01.1	PPB	4/26/2022 16:42	SMPL
N2264-13	YAZ01	648	0.1438 -		SFAM01.1	PPB	4/26/2022 16:44	SMPL
N2264-14	YAZ03	675	0.1498 -		SFAM01.1	PPB	4/26/2022 16:46	SMPL
PB144378BL	PBW378	17	0.0047 -		SFAM01.1	PPB	4/26/2022 16:51	SMPL
N2265-01	YAZ18	660	0.1465 -		SFAM01.1	PPB	4/26/2022 16:53	SMPL
N2265-02	YAZ18D	510	0.1134 -		SFAM01.1	PPB	4/26/2022 16:56	SMPL
N2265-03	YAZ18S	5228	1.1537 -		SFAM01.1	PPB	4/26/2022 17:01	SMPL
N2265-04	YAZ16	4424	0.9764 -		SFAM01.1	PPB	4/26/2022 17:03	SMPL
CCV037	CCV037	22776	5.0232 -		SFAM01.1	PPB	4/26/2022 17:05	SMPL
CCB037	CCB037	-51	-0.0103 -		SFAM01.1	PPB	4/26/2022 17:07	SMPL
N2265-05	YAZ22	643	0.1427 -		SFAM01.1	PPB	4/26/2022 17:10	SMPL
N2265-06	YAZ14	652	0.1447 -		SFAM01.1	PPB	4/26/2022 17:12	SMPL
N2265-07	YAYW2	428	0.0953 -		SFAM01.1	PPB	4/26/2022 17:14	SMPL
N2265-08	YAZ24	412	0.0918 -		SFAM01.1	PPB	4/26/2022 17:16	SMPL
N2265-09	YAZ66	375	0.0836 -		SFAM01.1	PPB	4/26/2022 17:19	SMPL
N2265-10	YAYW4	1278	0.2827 -		SFAM01.1	PPB	4/26/2022 17:21	SMPL
N2265-11	YAYW6	742	0.1645 -		SFAM01.1	PPB	4/26/2022 17:23	SMPL
N2265-12	YAZ00	998	0.221 -		SFAM01.1	PPB	4/26/2022 17:26	SMPL
N2265-13	YAZ02	610	0.1354 -		SFAM01.1	PPB	4/26/2022 17:28	SMPL
N2265-14	YAZ04	604	0.1341 -		SFAM01.1	PPB	4/26/2022 17:30	SMPL
CCV038	CCV038	23656	5.2172 -		SFAM01.1	PPB	4/26/2022 17:32	SMPL
CCB038	CCB038	-33	-0.0064 -		SFAM01.1	PPB	4/26/2022 17:35	SMPL