


LB 119853

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.3851e-004

C= -4.6219e-002

Rho= 0.9991939

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.033	-0.033	55	0.000	55					8
0.2	0.200	0.216	0.016	1098	0.0 %	1098					-2
2.5	2.500	2.444	-0.056	10439	0.0 %	10439					5
5.0	5.000	5.264	0.264	22265	0.0 %	22265					-3
7.5	7.500	7.265	-0.235	30653	0.0 %	30653					6
10.0	10.000	10.045	0.045	42310	0.0 %	42310					

LB119853 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Apr 2022 10:32:16

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	55	-	0.0007470A	PPB	27 Apr 2022 10:37:31	S	Std
0.2 - 1	50.2	1098	-	0.20007470A	PPB	27 Apr 2022 10:39:48	S	Std
2.5 - 1	52.5	10439	-	2.50007470A	PPB	27 Apr 2022 10:42:04	S	Std
5.0 - 1	55	22265	-	5.00007470A	PPB	27 Apr 2022 10:44:21	S	Std
7.5 - 1	57.5	30653	-	7.50007470A	PPB	27 Apr 2022 10:46:37	S	Std
10.0 - 1	51.8	42310	-	10.0007470A	PPB	27 Apr 2022 10:48:53	S	Std
ICV96 - 1	ICV96	15386	3.6234	-7470A	PPB	27 Apr 2022 10:51:35	U	SMPL
ICB96 - 1	ICB96	-43	-0.0565	-7470A	PPB	27 Apr 2022 10:53:50	U	SMPL
CCV39 - 1	CCV39	20874	4.9324	-7470A	PPB	27 Apr 2022 11:01:40	U	SMPL
CCB39 - 1	CCB39	-92	-0.0682	-7470A	PPB	27 Apr 2022 11:03:55	U	SMPL
CRA - 1	CRA	787	0.1415	-7470A	PPB	27 Apr 2022 11:08:40	U	SMPL
HighStd - 1	HighStd	41220	9.7850	-7470A	PPB	27 Apr 2022 11:10:55	U	SMPL
ChkStd - 1	ChkStd	27810	6.5867	-7470A	PPB	27 Apr 2022 11:13:10	U	SMPL
PB144412BL - 1	PBW	-87	-0.0670	-7470A	PPB	27 Apr 2022 11:15:26	U	SMPL
PB144412BS - 1	LCSW	16532	3.8968	-7470A	PPB	27 Apr 2022 11:20:21	U	SMPL
N2431-06 - 1	WC-15-2-1-GR	149	-0.0107	-7470A	PPB	27 Apr 2022 11:22:37	U	SMPL
N2557-01 - 1	SOIL-CUTTINGS	12	-0.0434	-7470A	PPB	27 Apr 2022 11:24:53	U	SMPL
N2559-02 - 1	MH-10S	56	-0.0329	-7470A	PPB	27 Apr 2022 11:27:10	U	SMPL
N2559-05 - 1	MH-10D	53	-0.0336	-7470A	PPB	27 Apr 2022 11:29:26	U	SMPL
N2559-05DUP - 1	MH-10DDUP	50	-0.0343	-7470A	PPB	27 Apr 2022 11:31:41	U	SMPL
CCV40 - 1	CCV40	22189	5.2460	-7470A	PPB	27 Apr 2022 11:36:32	U	SMPL
CCB40 - 1	CCB40	-74	-0.0639	-7470A	PPB	27 Apr 2022 11:38:46	U	SMPL
N2559-05MS - 1	MH-10DMS	14511	3.4148	-7470A	PPB	27 Apr 2022 11:41:01	U	SMPL
N2559-05MSD - 1	MH-10DMSD	14250	3.3525	-7470A	PPB	27 Apr 2022 11:43:16	U	SMPL
PB144413BL - 1	PBW	-25	-0.0522	-7470A	PPB	27 Apr 2022 11:45:32	U	SMPL
PB144413BS - 1	LCSW	15911	3.7487	-7470A	PPB	27 Apr 2022 11:50:22	U	SMPL
N2551-01 - 1	20220375-096-21-01-ELUTRIATE	-3	-0.0469	-7470A	PPB	27 Apr 2022 11:52:37	U	SMPL
N2551-01DUP - 1	20220375-096-21-01-ELUTRIATEDUP	54	-0.0333	-7470A	PPB	27 Apr 2022 11:54:53	U	SMPL
N2551-01MS - 1	20220375-096-21-01-ELUTRIATEMS	14797	3.4830	-7470A	PPB	27 Apr 2022 11:57:09	U	SMPL
N2551-01MSD - 1	20220375-096-21-01-ELUTRIATEMSD	14552	3.4245	-7470A	PPB	27 Apr 2022 11:59:26	U	SMPL
N2551-02 - 1	20220376-096-21-02-ELUTRIATE	31	-0.0388	-7470A	PPB	27 Apr 2022 12:01:43	U	SMPL
N2551-03 - 1	20220377-096-21-03-ELUTRIATE	90	-0.0248	-7470A	PPB	27 Apr 2022 12:03:59	U	SMPL
CCV41 - 1	CCV41	21537	5.0905	-7470A	PPB	27 Apr 2022 12:06:14	U	SMPL
CCB41 - 1	CCB41	-101	-0.0703	-7470A	PPB	27 Apr 2022 12:08:30	U	SMPL
N2551-04 - 1	20220378-098-21-01-ELUTRIATE	36	-0.0376	-7470A	PPB	27 Apr 2022 12:10:45	U	SMPL
N2557-03 - 1	PURGE-WATER	1222	0.2452	-7470A	PPB	27 Apr 2022 12:13:00	U	SMPL
PB144364TB - 1	PB144364TB	-47	-0.0574	-7470A	PPB	27 Apr 2022 12:15:16	U	SMPL
N2559-05LX5 - 1		26	-0.0400	-7470A	PPB	27 Apr 2022 12:17:31	U	SMPL
N2559-05A - 1		16360	3.8558	-7470A	PPB	27 Apr 2022 12:19:47	U	SMPL
N2551-01LX5 - 1		-24	-0.0519	-7470A	PPB	27 Apr 2022 12:22:03	U	SMPL
N2551-01A - 1		16210	3.8200	-7470A	PPB	27 Apr 2022 12:24:20	U	SMPL
CCV42 - 1	CCV42	21306	5.0354	-7470A	PPB	27 Apr 2022 12:26:37	U	SMPL
CCB42 - 1	CCB42	-83	-0.0660	-7470A	PPB	27 Apr 2022 12:28:54	U	SMPL