



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB119859

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	04/27/2022	09:33
2	CAL2	1	Water	NA	NA	20.4	7.00	04/27/2022	09:34
3	CAL3	1	Water	NA	NA	20.4	10.02	04/27/2022	09:35
4	ICV	1	Water	NA	NA	20.2	6.99	04/27/2022	09:36
5	CCV1	1	Water	NA	NA	20.5	2.01	04/27/2022	09:37
6	N2575-01	1	Water	NA	NA	20.5	7.47	04/27/2022	09:40
7	N2575-02	1	Water	NA	NA	20.6	6.68	04/27/2022	09:42
8	N2575-03	1	Water	NA	NA	20.7	7.29	04/27/2022	09:45
9	N2575-04	1	Water	NA	NA	20.6	6.85	04/27/2022	09:47
10	N2575-06	1	Water	NA	NA	20.2	6.67	04/27/2022	09:50
11	N2575-07	1	Water	NA	NA	20.5	6.49	04/27/2022	09:53
12	N2575-08	1	Water	NA	NA	20.7	6.37	04/27/2022	09:56
13	N2575-09	1	Water	NA	NA	20.2	6.42	04/27/2022	09:59
14	N2575-10	1	Water	NA	NA	20.2	6.80	04/27/2022	10:00
15	N2575-10DUP	1	Water	NA	NA	20.3	6.81	04/27/2022	10:01
16	CCV2	1	Water	NA	NA	20.4	12.03	04/27/2022	10:02

WB 714859

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2566 WorkList ID : 159043 Department : Wet-Chemistry Date : 04-27-2022 09:27:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/26/2022	Water	N2575-01	pH C	Cool 4 deg C	PSEG03	K11	31809	04/26/2022	9040C
04/26/2022	Water	N2575-02	pH C	Cool 4 deg C	PSEG03	K11	31799	04/26/2022	9040C
04/26/2022	Water	N2575-03	pH C	Cool 4 deg C	PSEG03	K11	31651	04/26/2022	9040C
04/26/2022	Water	N2575-04	pH C	Cool 4 deg C	PSEG03	K11	31652	04/26/2022	9040C
04/26/2022	Water	N2575-06	pH C	Cool 4 deg C	PSEG03	K11	COMP-95-98	04/26/2022	9040C
04/26/2022	Water	N2575-07	pH	Cool 4 deg C	PSEG03	K11	31795	04/26/2022	9040C
04/26/2022	Water	N2575-08	pH	Cool 4 deg C	PSEG03	K11	31796	04/26/2022	9040C
04/26/2022	Water	N2575-09	pH	Cool 4 deg C	PSEG03	K11	31797	04/26/2022	9040C
04/26/2022	Water	N2575-10	pH	Cool 4 deg C	PSEG03	K11	31798	04/26/2022	9040C

Date/Time 04/27/22 09:30
 Raw Sample Received by: JP (COC)
 Raw Sample Relinquished by: RS (CEA-106)

Date/Time 04/27/22 11:30
 Raw Sample Received by: RS (CEA-106)
 Raw Sample Relinquished by: JP (COC)