



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB119860

Slope : 98.6

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/27/2022	13:00
2	CAL2	1	Water	NA	NA	20.2	7.01	04/27/2022	13:01
3	CAL3	1	Water	NA	NA	20.4	10.03	04/27/2022	13:03
4	ICV	1	Water	NA	NA	20.5	6.99	04/27/2022	13:05
5	CCV1	1	Water	NA	NA	20.3	2.02	04/27/2022	13:06
6	N2572-02	1	Solid	20.03	20	21.2	7.10	04/27/2022	13:10
7	N2572-05	1	Solid	20.04	20	20.7	7.00	04/27/2022	13:11
8	N2588-02	1	Solid	20.02	20	20.9	8.43	04/27/2022	13:15
9	N2588-04	1	Solid	20.03	20	21.1	9.01	04/27/2022	13:19
10	N2588-04DUP	1	Solid	20.04	20	21.2	9.02	04/27/2022	13:20
11	CCV2	1	Water	NA	NA	20.9	12.03	04/27/2022	13:24

VP 119860

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2577 WorkList ID : 159060 Department : Wet-Chemistry Date : 04-27-2022 12:44:28

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
04/26/2022	Solid	N2572-02	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-14S	04/26/2022	9045D
04/26/2022	Solid	N2572-05	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-14D	04/26/2022	9045D
04/27/2022	Solid	N2588-02	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-1	04/27/2022	9045D
04/27/2022	Solid	N2588-04	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-2	04/27/2022	9045D

Date/Time 04/27/22 12:50
Raw Sample Received by: JP (GRI)
Raw Sample Relinquished by: JP (GRI)

Date/Time 04/27/22 14:30
Raw Sample Received by: JP (GRI)
Raw Sample Relinquished by: JP (GRI)