


LB 119861

245.1

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.7401e-004

C= -3.4327e-001

Rho= 0.9994541

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.107	-0.107	864	0.000	864					-6
0.05	0.050	0.047	-0.003	1423	0.0 %	1423	0				-5
0.2	0.200	0.191	-0.009	1949	0.0 %	1949					5
2.5	2.500	2.619	0.119	10810	0.0 %	10810					-1
5.0	5.000	4.975	-0.025	19409	0.0 %	19409					3
7.5	7.500	7.711	0.211	29394	0.0 %	29394					-2
10.0	10.000	9.814	-0.186	37069	0.0 %	37069					

LB119861 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Apr 2022 14:39:54

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	864	-	0.0000	245.1	PPB	27 Apr 2022 14:55:04	S	Std
0.05 - 1	50.05	1423	-	0.0500	245.1	PPB	27 Apr 2022 14:57:21	S	Std
0.2 - 1	50.2	1949	-	0.2000	245.1	PPB	27 Apr 2022 14:59:38	S	Std
2.5 - 1	52.5	10810	-	2.5000	245.1	PPB	27 Apr 2022 15:01:55	S	Std
5.0 - 1	55	19409	-	5.0000	245.1	PPB	27 Apr 2022 15:04:11	S	Std
7.5 - 1	57.5	29394	-	7.5000	245.1	PPB	27 Apr 2022 15:06:27	S	Std
10.0 - 1	511	37069	-	10.0000	245.1	PPB	27 Apr 2022 15:13:51	S	Std
ICV97 - 1	ICV97	15784	3.9818	-245.1		PPB	27 Apr 2022 15:16:47	U	SMPL
ICB97 - 1	ICB97	838	-0.1136	-245.1		PPB	27 Apr 2022 15:21:34	U	SMPL
CCV43 - 1	CCV43	19662	5.0444	-245.1		PPB	27 Apr 2022 15:23:48	U	SMPL
CCB43 - 1	CCB43	814	-0.1202	-245.1		PPB	27 Apr 2022 15:28:13	U	SMPL
CRA - 1	CRA	1388	0.0371	-245.1		PPB	27 Apr 2022 15:42:48	U	SMPL
HighStd - 1	HighStd	36487	9.6547	-245.1		PPB	27 Apr 2022 15:45:04	U	SMPL
ChkStd - 1	ChkStd	26258	6.8518	-245.1		PPB	27 Apr 2022 15:47:19	U	SMPL
PB144426BL - 1	PBW	992	-0.0714	-245.1		PPB	27 Apr 2022 15:52:09	U	SMPL
PB144426BS - 1	LCSW	15431	3.8850	-245.1		PPB	27 Apr 2022 15:56:56	U	SMPL
N2550-01 - 1	PUMP-STATION-207ST-TRAIN-FACILITY	1248	-0.0013	-245.1		PPB	27 Apr 2022 15:59:12	U	SMPL
N2550-01DUP - 1	PUMP-STATION-207ST-TRAIN-FACILITYDUP	625	-0.1720	-245.1		PPB	27 Apr 2022 16:01:29	U	SMPL
N2550-01MS - 1	PUMP-STATION-207ST-TRAIN-FACILITYMS	15349	3.8626	-245.1		PPB	27 Apr 2022 16:03:45	U	SMPL
N2550-01MSD - 1	PUMP-STATION-207ST-TRAIN-FACILITYMSD	15142	3.8058	-245.1		PPB	27 Apr 2022 16:06:02	U	SMPL
N2550-02 - 1	DYCKMAN-200ST-A51515T50SB	941	-0.0854	-245.1		PPB	27 Apr 2022 16:10:31	U	SMPL
CCV44 - 1	CCV44	19125	4.8972	-245.1		PPB	27 Apr 2022 16:12:46	U	SMPL
CCB44 - 1	CCB44	919	-0.0914	-245.1		PPB	27 Apr 2022 16:17:33	U	SMPL
N2550-01LX5 - 1		1417	0.0450	-245.1		PPB	27 Apr 2022 16:22:22	U	SMPL
N2550-01A - 1		15162	3.8113	-245.1		PPB	27 Apr 2022 16:24:37	U	SMPL
CCV45 - 1	CCV45	19033	4.8720	-245.1		PPB	27 Apr 2022 16:26:52	U	SMPL
CCB45 - 1	CCB45	863	-0.1068	-245.1		PPB	27 Apr 2022 16:29:08	U	SMPL