



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/28/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 15:35
In Date: 04/27/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:12
Out Date: 04/28/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119863

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2579-01	GBEL1	1	1.13	8.58	9.71	9.58	98.5	
N2579-02	GBEL2	2	1.13	8.85	9.98	9.88	98.9	
N2579-03	GBEL3	3	1.12	8.42	9.54	9.5	99.5	
N2579-04	GBEL4	4	1.13	8.85	9.98	9.93	99.4	
N2579-05	GBEL5	5	1.12	8.48	9.6	9.48	98.6	
N2579-06	GBEL6	6	1.14	8.48	9.62	9.5	98.6	
N2579-07	GBEL7	7	1.13	8.52	9.65	9.56	98.9	
N2579-08	GBEL8	8	1.14	8.72	9.86	9.8	99.3	
N2579-09	GBEL9	9	1.14	8.64	9.78	9.68	98.8	
N2579-10	GBEM0	10	1.15	8.70	9.85	9.7	98.3	
N2579-11	GBEM1	11	1.13	8.76	9.89	9.8	99.0	
N2579-12	GBEM2	12	1.13	8.68	9.81	9.66	98.3	
N2579-13	GBEM3	13	1.12	8.66	9.78	9.7	99.1	
N2579-14	GBEM4	14	1.13	8.43	9.56	9.52	99.5	
N2579-15	GBEM4D	15	1.13	8.43	9.56	9.52	99.5	
N2579-16	GBEM4S	16	1.13	8.43	9.56	9.52	99.5	
N2579-17	GBEM5	17	1.12	8.53	9.65	9.57	99.1	
N2579-18	GBEM6	18	1.13	8.47	9.6	9.54	99.3	
N2579-19	GBEM7	19	1.13	8.71	9.84	9.77	99.2	
N2579-20	GBEM8	20	1.12	8.65	9.77	9.7	99.2	
N2579-21	GBEM9	21	1.13	8.85	9.98	9.91	99.2	
N2579-22	GBEN0	22	1.16	8.79	9.95	9.92	99.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$