



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB119868

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/27/2022	13:20
2	N2575-01		1	15	100.00	>212.0	>212.0	04/27/2022	13:45
3	N2575-01DUP		1	15	100.00	>212.0	>212.0	04/27/2022	14:10
4	N2575-02		1	12	100.00	>212.0	>212.0	04/27/2022	14:35
5	N2575-03		1	15	100.00	>212.0	>212.0	04/27/2022	14:58
6	N2575-04		1	14	100.00	>212.0	>212.0	04/27/2022	15:25
7	N2575-06		1	10	100.00	>212.0	>212.0	04/27/2022	15:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB119868

WorkList Name : FP42722 WorkList ID : 159057 Department : Wet-Chemistry Date : 04-27-2022 11:20:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Customer Sample	Collect Date	Method
						Storage	Location			
05/03/2022	Water	N2575-01	Flash Point	Cool 4 deg C	PSEG03	K11	31809		04/26/2022	1010B
05/03/2022	Water	N2575-02	Flash Point	Cool 4 deg C	PSEG03	K11	31799		04/26/2022	1010B
05/03/2022	Water	N2575-03	Flash Point	Cool 4 deg C	PSEG03	K11	31651		04/26/2022	1010B
05/03/2022	Water	N2575-04	Flash Point	Cool 4 deg C	PSEG03	K11	31652		04/26/2022	1010B
05/03/2022	Water	N2575-06	Flash Point	Cool 4 deg C	PSEG03	K11	COMP-95-98		04/26/2022	1010B

Date/Time 04/27/22 12:55
 Raw Sample Received by: KS (wc)
 Raw Sample Relinquished by: RJ C17A-106b)

Date/Time 04/27/22 15:55
 Raw Sample Received by: RJ C17A-106b)
 Raw Sample Relinquished by: KS (wc)