

Clear table

Instrument ID: IC-2 Analyst: IZ Method: 300.0 / 9056A

ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-04-25_	4/25/22 10:49		IZ/AP
STD2	0.4940	0.6600	0.6580	2.1980	0.5550	0.8440	3.4480	2022-04-25_	4/25/22 11:18		IZ/AP
STD3	0.7750	1.2070	1.2080	4.0320	1.0050	2.0310	6.0790	2022-04-25_	4/25/22 11:46		IZ/AP
STD4	0.9970	1.4640	1.4570	4.8770	1.2260	2.5780	7.2990	2022-04-25_	4/25/22 12:15		IZ/AP
STD5	1.9260	2.9560	2.9650	9.8570	2.4560	5.0880	14.5850	2022-04-25_	4/25/22 12:44		IZ/AP
STD6	3.9310	5.9740	5.9840	19.8990	4.9610	10.0330	29.7380	2022-04-25_	4/25/22 13:12		IZ/AP
STD7	5.0770	7.5390	7.5290	25.1380	6.2970	12.4270	37.8510	2022-04-25_	4/25/22 13:41		IZ/AP
ICV	2.0000	2.8960	2.8920	9.6520	2.4140	5.4740	14.2740	2022-04-25_	4/25/22 14:12		IZ/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-04-25_	4/25/22 14:41		IZ/AP
CCV	1.9430	2.8670	2.8710	9.5500	2.3660	4.8870	14.1030	2022-04-27_	4/27/22 9:44		IZ/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-04-27_	4/27/22 10:13		IZ/AP
LB119871BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-04-27_	4/27/22 10:42		IZ/AP
LB119871BSW	1.9130	2.8660	2.8830	9.5550	2.3770	5.1100	14.1210	2022-04-27_	4/27/22 11:11		IZ/AP
N2593-01X10	0.0000	101.2940	0.0000	0.7970	0.7940	0.0000	5.2500	2022-04-27_	4/27/22 16:32		IZ/AP
N2593-01DLX300	0.0000	2.4790	0.0000	0.0000	0.0000	0.0000	0.0000	2022-04-27_	4/27/22 17:01		IZ/AP
N2593-01MSX10	0.2690	98.8890	0.3130	1.7170	1.0100	1.2190	6.5130	2022-04-27_	4/27/22 17:29		IZ/AP
N2593-01MSDX10	0.2800	97.3040	0.3060	1.7040	1.0100	1.3010	6.4320	2022-04-27_	4/27/22 17:58		IZ/AP
CCV	1.9340	2.9010	2.9080	9.6460	2.4020	5.7770	14.2210	2022-04-27_	4/27/22 18:27		IZ/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-04-27_	4/27/22 18:55		IZ/AP

Clear table

Instrument ID: IC-2 Analyst: IZ Method: 300.0 / 9056A


ident	concentratio tion F-	concentratio n CL-	concentratio on NO2	concentratio on BR-	concentratio on NO3	concentratio on HPO4	concentratio on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-04-25_	4/25/22 10:49		IZ/AP
STD2	0.4940	0.6600	0.6580	2.1980	0.5550	0.8440	3.4480	2022-04-25_	4/25/22 11:18		IZ/AP
STD3	0.7750	1.2070	1.2080	4.0320	1.0050	2.0310	6.0790	2022-04-25_	4/25/22 11:46		IZ/AP
STD4	0.9970	1.4640	1.4570	4.8770	1.2260	2.5780	7.2990	2022-04-25_	4/25/22 12:15		IZ/AP
STD5	1.9260	2.9560	2.9650	9.8570	2.4560	5.0880	14.5850	2022-04-25_	4/25/22 12:44		IZ/AP
STD6	3.9310	5.9740	5.9840	19.8990	4.9610	10.0330	29.7380	2022-04-25_	4/25/22 13:12		IZ/AP
STD7	5.0770	7.5390	7.5290	25.1380	6.2970	12.4270	37.8510	2022-04-25_	4/25/22 13:41		IZ/AP

ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000

ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4
STD1							
STD2	23.5000	10.0000	9.6667	9.9000	11.0000	-15.6000	14.9333
STD3	-3.1250	0.5833	0.6667	0.8000	0.5000	1.5500	1.3167
STD4	-0.3000	-2.4000	-2.8667	-2.4600	-1.9200	3.1200	-2.6800
STD5	-3.7000	-1.4667	-1.1667	-1.4300	-1.7600	1.7600	-2.7667
STD6	-1.7250	-0.4333	-0.2667	-0.5050	-0.7800	0.3300	-0.8733
STD7	1.5400	0.5200	0.3867	0.5520	0.7520	-0.5840	2.3000

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-042522.mtw
Equation: $Q = 0.0784305 \cdot A + 2.35146$
RSD: 3.637 %
Correlation coefficient: 0.999279




K3 = 0 K2 = 0 K1 = 0.0784305 K0 = 2.35146
Base: Area
Ref.channel: ch1
ISTD:
Formula: Linear
Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	1.965	95.99	0.4	20	5.653		
3	4.019	167.7	0.8	20	5.653		
4	5.79	224.3	1	20	5.653		
5	11.46	461.1	2	20	5.653		
6	25.77	972.4	4	20	5.653		
7	35.65	1265	5	20	5.653		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-042522.mtw
 Equation: $Q = 0.119071 \cdot A + 2.57053$
 RSD: 1.448 %
 Correlation coefficient: 0.999886




K3 = 0 K2 = 0 K1 = 0.119071 K0 = 2.57053
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.071	89.31	0.6	20	8.643		
3	12.35	181.2	1.2	20	8.643		
4	15.33	224.4	1.5	20	8.643		
5	32.79	474.9	3	20	8.643		
6	68.91	981.8	6	20	8.643		
7	87.36	1245	7.5	20	8.643		

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-042522.mtw
 Equation: $Q = 0.0523088 \cdot A + 2.71284$
 RSD: 1.317 %
 Correlation coefficient: 0.999905




K3 = 0 K2 = 0 K1 = 0.0523088 K0 = 2.71284
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0		0		0
2	11.26	199.6		0.6	20		10.45
3	22.68	410		1.2	20		10.45
4	28.41	505.2		1.5	20		10.45
5	59.63	1082		3	20		10.45
6	117.1	2236		6	20		10.45
7	147.6	2827		7.5	20		10.45

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-042522.mtw
 Equation: $Q = 0.266226 \cdot A + 9.23936$
 RSD: 1.471 %
 Correlation coefficient: 0.999882




K3 = 0 K2 = 0 K1 = 0.266226 K0 = 9.23936
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.214	130.4	2	20			13.53
3	12.91	268.2	4	20			13.53
4	16.03	331.7	5	20			13.53
5	34.52	705.8	10	20			13.53
6	71.72	1460	20	20			13.53
7	90.62	1854	25	20			13.53

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-042522.mtw
 Equation: $Q = 0.0464515 \cdot A + 2.4972$
 RSD: 1.753 %
 Correlation coefficient: 0.999833



K3 = 0 K2 = 0 K1 = 0.0464515 K0 = 2.4972

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.427	185.2	0.5	20	15.58		
3	15.4	379.1	1	20	15.58		
4	19.24	474	1.25	20	15.58		
5	41.09	1004	2.5	20	15.58		
6	86.14	2082	5	20	15.58		
7	110	2657	6.25	20	15.58		

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-042522.mtw
 Equation: $Q = 0.0858704 \cdot A - 1.26178$
 RSD: 1.938 %
 Correlation coefficient: 0.999795




K3 = 0 K2 = 0 K1 = 0.0858704 K0 = -1.26178
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	5.391	211.2	1	20	20.18		
3	13.17	487.7	2	20	20.18		
4	16.71	615.1	2.5	20	20.18		
5	33.74	1200	5	20	20.18		
6	70.69	2351	10	20	20.18		
7	88.83	2909	12.5	20	20.18		

±□;□

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-042522.mtw
 Equation: $Q = 0.150846 \cdot A + 16.9558$
 RSD: 2.370 %
 Correlation coefficient: 0.999694



K3 = 0 K2 = 0 K1 = 0.150846 K0 = 16.9558

Base: Area

Ref.channel: chl

ISTD:

Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	11.17	344.8	3	20	23.09		
3	22.74	693.6	6	20	23.09		
4	28.22	855.4	7.5	20	23.09		
5	61.17	1821	15	20	23.09		
6	130.6	3830	30	20	23.09		
7	167.7	4906	37.5	20	23.09		

Report date: 4/25/2022 2:11:52 PM
Printed by: wet


Ident: STD1
Analysis from: 4/25/2022 10:49:32 AM
File: _2022-04-25_

Last save: 4/25/2022 2:10:58 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30128

Last save: 4/25/2022 1:38:

SAMPLE:
: IZ/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/25/2022 2:12:02 PM
Printed by: wet


Ident: STD2
Analysis from: 4/25/2022 11:18:09 AM
File: _2022-04-25_

Last save: 4/25/2022 2:10:58 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30129

Last save: 4/25/2022 1:38:

SAMPLE:
: IZ/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.588	1.96	3.97	95.985	7.64
2	8.84	0.206	6.07	12.26	89.314	7.11
3	10.68	0.242	11.26	22.74	199.628	15.89
4	13.85	0.304	6.21	12.55	130.385	10.38
5	15.95	0.358	7.43	15.01	185.180	14.74
6	20.51	0.510	5.39	10.89	211.243	16.81
7	23.88	0.457	11.17	22.57	344.759	27.44
7	26.00	0.381	49.49	99.99	1256.493	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/25/2022 2:12:09 PM
Printed by: wet


Ident: STD3
Analysis from: 4/25/2022 11:46:48 AM
File: _2022-04-25_

Last save: 4/25/2022 2:10:58 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30130

Last save: 4/25/2022 1:38:

SAMPLE:
: IZ/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.518	4.02	3.89	167.718	6.48
2	8.82	0.204	12.35	11.96	181.210	7.00
3	10.65	0.246	22.68	21.96	409.952	15.84
4	13.78	0.300	12.91	12.50	268.176	10.36
5	15.85	0.354	15.40	14.91	379.103	14.65
6	20.50	0.488	13.17	12.75	487.695	18.85
7	23.89	0.452	22.73	22.02	693.573	26.81
7	26.00	0.366	103.26	100.00	2587.426	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/25/2022 2:12:16 PM
Printed by: wet


Ident: STD4
Analysis from: 4/25/2022 12:15:25 PM
File: _2022-04-25_

Last save: 4/25/2022 2:10:58 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30131

Last save: 4/25/2022 1:38:

SAMPLE:
: IZ/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.84	0.480	5.79	4.46	224.308	6.94
2	8.83	0.204	15.33	11.82	224.383	6.95
3	10.66	0.243	28.41	21.90	505.188	15.64
4	13.79	0.299	16.03	12.36	331.688	10.27
5	15.85	0.352	19.24	14.83	473.950	14.67
6	20.55	0.483	16.71	12.88	615.063	19.04
7	23.94	0.450	28.22	21.75	855.363	26.48
7	26.00	0.359	129.73	99.99	3229.944	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/25/2022 2:12:23 PM
Printed by: wet


Ident: STD5
Analysis from: 4/25/2022 12:44:03 PM
File: _2022-04-25_

Last save: 4/25/2022 2:10:58 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30132

Last save: 4/25/2022 1:38:

SAMPLE:
: IZ/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.81	0.500	11.45	4.17	461.094	6.83
2	8.79	0.201	32.79	11.95	474.851	7.04
3	10.59	0.247	59.62	21.73	1081.752	16.03
4	13.69	0.294	34.52	12.58	705.807	10.46
5	15.71	0.348	41.09	14.98	1003.790	14.87
6	20.45	0.470	33.74	12.30	1199.670	17.78
7	23.84	0.441	61.17	22.29	1821.329	26.99
7	26.00	0.357	274.40	100.00	6748.293	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/25/2022 2:12:30 PM
Printed by: wet


Ident: STD6
Analysis from: 4/25/2022 1:12:41 PM
File: _2022-04-25_

Last save: 4/25/2022 2:10:58 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30133

Last save: 4/25/2022 1:38:

SAMPLE:
: IZ/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.476	25.77	4.51	972.439	6.99
2	8.80	0.197	68.91	12.07	981.766	7.06
3	10.58	0.262	117.12	20.51	2236.083	16.07
4	13.70	0.293	71.72	12.56	1460.160	10.49
5	15.68	0.343	86.13	15.09	2082.165	14.96
6	20.47	0.446	70.69	12.38	2351.430	16.90
7	23.87	0.434	130.61	22.88	3830.400	27.53
7	26.00	0.350	570.96	100.00	13914.442	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/25/2022 2:12:38 PM
Printed by: wet


Ident: STD7
Analysis from: 4/25/2022 1:41:19 PM
File: _2022-04-25_

Last save: 4/25/2022 2:10:58 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30134

Last save: 4/25/2022 1:38:

SAMPLE:
: IZ/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.84	0.456	35.65	4.90	1264.605	7.16
2	8.81	0.197	87.35	12.00	1244.686	7.05
3	10.58	0.263	147.55	20.28	2826.654	16.00
4	13.71	0.295	90.62	12.45	1853.743	10.50
5	15.67	0.342	109.98	15.11	2657.434	15.05
6	20.51	0.441	88.83	12.21	2909.061	16.47
7	23.92	0.434	167.74	23.05	4906.119	27.78
7	26.00	0.347	727.72	100.00	17662.302	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/25/2022 2:41:40 PM
Printed by: wet


Ident: ICV
Analysis from: 4/25/2022 2:12:06 PM
File: _2022-04-25_

Last save: 4/25/2022 2:38:07 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30135

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.507	11.80	4.31	480.046	7.12
2	8.78	0.200	32.29	11.80	464.877	6.89
3	10.58	0.245	58.45	21.37	1054.012	15.63
4	13.69	0.292	34.19	12.50	690.418	10.24
5	15.69	0.345	40.60	14.84	985.566	14.61
6	20.47	0.470	36.10	13.20	1289.668	19.12
7	23.85	0.440	60.14	21.98	1780.183	26.39
7	26.00	0.357	273.56	100.00	6744.769	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/25/2022 3:08:05 PM
Printed by: wet


Ident: ICB
Analysis from: 4/25/2022 2:41:30 PM
File: _2022-04-25_

Last save: 4/25/2022 3:07:30 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30136

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/28/2022 9:33:14 AM
Printed by: wet


Ident: CCV
Analysis from: 4/27/2022 9:44:37 AM
File: _2022-04-27_

Last save: 4/27/2022 10:10:37 AM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30155

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.537	10.99	4.17	465.605	7.13
2	8.82	0.201	31.73	12.05	460.003	7.04
3	10.63	0.247	57.27	21.74	1045.855	16.02
4	13.77	0.296	33.34	12.66	682.706	10.46
5	15.81	0.350	39.51	15.00	965.089	14.78
6	20.48	0.477	31.92	12.12	1152.839	17.66
7	23.90	0.445	58.67	22.27	1757.425	26.92
7	26.00	0.365	263.42	100.00	6529.522	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/28/2022 9:33:25 AM
Printed by: wet


Ident: CCB
Analysis from: 4/27/2022 10:13:59 AM
File: _2022-04-27_

Last save: 4/27/2022 10:39:59 AM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30156

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/28/2022 9:33:35 AM
Printed by: wet


Ident: LB119871BLW
Analysis from: 4/27/2022 10:42:37 AM
File: _2022-04-27_

Last save: 4/28/2022 9:04:57 AM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30157

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 4/28/2022 9:33:43 AM
Printed by: wet


Ident: LB119871BSW
Analysis from: 4/27/2022 11:11:15 AM
File: _2022-04-27_

Last save: 4/28/2022 9:04:57 AM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30158

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.81	0.510	11.23	4.16	457.775	6.95
2	8.76	0.199	32.16	11.92	459.732	6.98
3	10.54	0.244	58.52	21.69	1050.300	15.95
4	13.63	0.290	34.13	12.65	683.138	10.37
5	15.62	0.342	40.58	15.04	969.510	14.72
6	20.44	0.471	33.84	12.54	1204.890	18.30
7	23.85	0.441	59.31	21.98	1759.824	26.72
7	26.00	0.357	269.77	100.00	6585.169	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/28/2022 9:33:58 AM
Printed by: wet


Ident: N2593-01X10
Analysis from: 4/27/2022 4:32:38 PM
File: _2022-04-27_

Last save: 4/27/2022 4:58:39 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30160

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.92	0.210	1126.66	97.30	16992.373	94.98
3	0.00	0.000	0.00	0.00	0.000	0.00
4	14.11	0.297	1.28	0.11	25.195	0.14
5	16.14	0.356	11.74	1.01	288.295	1.61
6	0.00	0.000	0.00	0.00	0.000	0.00
7	24.08	0.470	18.31	1.58	583.684	3.26
7	26.00	0.191	1157.99	100.00	17889.547	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/28/2022 9:35:19 AM
Printed by: wet


Ident: N2593-01DLX300
Analysis from: 4/27/2022 5:01:16 PM
File: _2022-04-27_

Last save: 4/28/2022 9:35:15 AM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30161

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.82	0.206	26.58	99.99	394.880	100.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	26.00	0.029	26.58	99.99	394.880	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/28/2022 9:35:34 AM
Printed by: wet


Ident: N2593-01MSX10
Analysis from: 4/27/2022 5:29:54 PM
File: _2022-04-27_

Last save: 4/27/2022 5:55:54 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30162

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.530	1.15	0.10	38.559	0.21
2	8.84	0.208	1114.39	95.19	16588.421	91.05
3	10.68	0.257	3.97	0.34	67.789	0.37
4	13.92	0.290	4.81	0.41	94.310	0.52
5	15.93	0.349	15.87	1.36	381.147	2.09
6	20.59	0.581	6.55	0.56	298.702	1.64
7	23.96	0.465	23.96	2.05	751.061	4.12
7	26.00	0.383	1170.70	100.00	18219.988	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/28/2022 9:35:46 AM
Printed by: wet


Ident: N2593-01MSDX10
Analysis from: 4/27/2022 5:58:34 PM
File: _2022-04-27_

Last save: 4/27/2022 6:24:34 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30163

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 18
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.85	0.538	1.23	0.11	41.346	0.23
2	8.86	0.208	1089.03	95.13	16322.238	90.87
3	10.69	0.253	3.90	0.34	65.234	0.36
4	13.93	0.292	4.70	0.41	93.340	0.52
5	15.95	0.352	15.57	1.36	380.918	2.12
6	20.64	0.591	6.82	0.60	317.823	1.77
7	24.02	0.468	23.47	2.05	740.365	4.12
7	26.00	0.386	1144.73	100.00	17961.264	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/28/2022 9:35:54 AM
Printed by: wet


Ident: CCV
Analysis from: 4/27/2022 6:27:16 PM
File: _2022-04-27_

Last save: 4/27/2022 6:53:16 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30164

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.81	0.514	11.32	4.24	463.252	6.82
2	8.78	0.204	31.72	11.89	465.607	6.85
3	10.58	0.252	57.32	21.50	1060.095	15.61
4	13.69	0.300	33.27	12.48	689.969	10.16
5	15.72	0.356	39.43	14.79	980.636	14.44
6	20.43	0.494	36.38	13.64	1360.162	20.02
7	23.84	0.460	57.23	21.46	1773.114	26.10
7	26.00	0.369	266.66	100.00	6792.835	100.00

This report has been created by IC Net
METROHM LTD

Report date: 4/28/2022 9:36:03 AM
Printed by: wet


Ident: CCB
Analysis from: 4/27/2022 6:55:53 PM
File: _2022-04-27_

Last save: 4/27/2022 7:21:53 PM

Method: AnionsIC2-042522.mtw
Run operator: wet
Analysis number: 30165

Last save: 4/25/2022 2:07:

SAMPLE:
: IZ/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD