



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB119874

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP95371
COD calibration std. 100 ppm	WP95373
COD calibration std. 10 ppm	WP95375
COD calibration std. 150 ppm	WP95372
COD calibration std. 50 ppm	WP95374
COD calibration std. 0 ppm	WP95376
COD ICV-LCS std, 50ppm	WP96424
COD CCV std, 50ppm	WP96423
COD Digestion Vials Low Level 0-150Mg/L	W2830

Temp In (C): <u>148</u>	Date In: <u>04/27/2022</u>	Time In: <u>14:20</u>
Temp Out (C): <u>150</u>	Date Out: <u>04/27/2022</u>	Time Out: <u>16:20</u>

Intercept: -0.4433

Slope: 1.0136

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	0.437		02/18/2022	13:20
2	CAL2	10	1	Water	9.000	9.317	6.8	02/18/2022	13:20
3	CAL3	50	1	Water	51.000	50.753	-1.5	02/18/2022	13:21
4	CAL4	100	1	Water	100.000	99.096	0.9	02/18/2022	13:21
5	CAL5	150	1	Water	152.000	150.398	-0.3	02/18/2022	13:22



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB119874

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	49.000	48.780	02/18/2022	13:22
2	ICB		NA	NA	1	Water	0.000	0.437	02/18/2022	13:23
3	CCV1	50	NA	NA	1	Water	51.000	50.753	04/28/2022	10:05
4	CCB1		NA	NA	1	Water	1.000	1.424	04/28/2022	10:05
5	LB119874BL		NA	NA	1	Water	1.000	1.424	04/28/2022	10:06
6	LB119874BS	50	NA	NA	1	Water	50.000	49.766	04/28/2022	10:06
7	N2452-15		NA	NA	1	Water	52.000	51.740	04/28/2022	10:07
8	N2452-15DUP		NA	NA	1	Water	51.000	50.753	04/28/2022	10:07
9	N2452-15MS	50	NA	NA	1	Water	99.000	98.109	04/28/2022	10:08
10	N2452-15MSD	50	NA	NA	1	Water	98.000	97.122	04/28/2022	10:08
11	CCV2	50	NA	NA	1	Water	49.000	48.780	04/28/2022	10:09
12	CCB2		NA	NA	1	Water	0.000	0.437	04/28/2022	10:09