


LB119875

7470A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.5512e-004

C= 1.0882e-002

Rho= 0.9996484

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.011	0.011	2	0.000	2					
0.2	0.200	0.246	0.046	921	0.0 %	921					23
2.5	2.500	2.364	-0.136	9224	0.0 %	9224					-5
5.0	5.000	5.156	0.156	20168	0.0 %	20168					3
7.5	7.500	7.393	-0.107	28935	0.0 %	28935					-1
10.0	10.000	10.030	0.030	39270	0.0 %	39270					0

LB119875 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Apr 2022 10:59:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	2	-	0.0000	7470A	PPB	28 Apr 2022 11:12:41	S	Std
0.2 - 1	50.2	921	-	0.2000	7470A	PPB	28 Apr 2022 11:14:58	S	Std
2.5 - 1	52.5	9224	-	2.5000	7470A	PPB	28 Apr 2022 11:17:15	S	Std
5.0 - 1	55	20168	-	5.0000	7470A	PPB	28 Apr 2022 11:19:32	S	Std
7.5 - 1	57.5	28935	-	7.5000	7470A	PPB	28 Apr 2022 11:21:48	S	Std
10.0 - 1	59	39270	-	10.0000	7470A	PPB	28 Apr 2022 11:26:34	S	Std
ICV98 - 1	ICV98	16267	4.1610	-7470A	PPB		28 Apr 2022 11:30:47	U	SMPL
ICB98 - 1	ICB98	-88	-0.0116	-7470A	PPB		28 Apr 2022 11:33:02	U	SMPL
CCV46 - 1	CCV46	20132	5.1470	-7470A	PPB		28 Apr 2022 11:35:17	U	SMPL
CCB46 - 1	CCB46	-82	-0.0100	-7470A	PPB		28 Apr 2022 11:37:31	U	SMPL
CRA - 1	CRA	900	0.2405	-7470A	PPB		28 Apr 2022 11:39:46	U	SMPL
HighStd - 1	HighStd	38031	9.7135	-7470A	PPB		28 Apr 2022 11:42:00	U	SMPL
ChkStd - 1	ChkStd	25994	6.6425	-7470A	PPB		28 Apr 2022 11:44:16	U	SMPL
PB144425BL - 1	PBW	-133	-0.0230	-7470A	PPB		28 Apr 2022 11:46:32	U	SMPL
PB144425BS - 1	LCSW	17044	4.3592	-7470A	PPB		28 Apr 2022 11:48:49	U	SMPL
N2560-02 - 1	2827	-30	0.0032	-7470A	PPB		28 Apr 2022 11:51:05	U	SMPL
N2564-02 - 1	PSC-164718	60	0.0262	-7470A	PPB		28 Apr 2022 11:53:21	U	SMPL
N2564-04 - 1	PSC-124019	32	0.0190	-7470A	PPB		28 Apr 2022 11:55:38	U	SMPL
N2572-02 - 1	MH-14S	23	0.0168	-7470A	PPB		28 Apr 2022 11:57:55	U	SMPL
N2572-05 - 1	MH-14D	25	0.0173	-7470A	PPB		28 Apr 2022 12:00:10	U	SMPL
CCV47 - 1	CCV47	19932	5.0960	-7470A	PPB		28 Apr 2022 12:02:26	U	SMPL
CCB47 - 1	CCB47	-60	-0.0044	-7470A	PPB		28 Apr 2022 12:04:41	U	SMPL
N2572-05DUP - 1	MH-14DDUP	12	0.0139	-7470A	PPB		28 Apr 2022 12:06:56	U	SMPL
N2572-05MS - 1	MH-14DMS	15213	3.8921	-7470A	PPB		28 Apr 2022 12:09:12	U	SMPL
N2572-05MSD - 1	MH-14DMSD	13639	3.4905	-7470A	PPB		28 Apr 2022 12:11:27	U	SMPL
PB144397TB - 1	PB144397TB	-55	-0.0031	-7470A	PPB		28 Apr 2022 12:13:42	U	SMPL
N2572-05LX5 - 1		15	0.0147	-7470A	PPB		28 Apr 2022 12:15:58	U	SMPL
N2572-05A - 1		15720	4.0214	-7470A	PPB		28 Apr 2022 12:18:14	U	SMPL
CCV48 - 1	CCV48	19744	5.0480	-7470A	PPB		28 Apr 2022 12:20:30	U	SMPL
CCB48 - 1	CCB48	-97	-0.0139	-7470A	PPB		28 Apr 2022 12:22:47	U	SMPL