



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/29/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 15:15
In Date: 04/28/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 04/29/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119877

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2585-01	GBF02	1	1.1	8.52	9.62	9.47	98.2	
N2585-02	GBF03	2	1.1	8.43	9.53	9.1	94.9	
N2585-03	GBF04	3	1.00	8.52	9.52	9.24	96.7	
N2585-04	GBF05	4	1.11	8.57	9.68	9.49	97.8	
N2585-05	GBF06	5	1.1	8.76	9.86	9.7	98.2	
N2585-06	GBF07	6	1.1	8.60	9.7	9.47	97.3	
N2585-07	GBF08	7	1.12	8.68	9.8	9.66	98.4	
N2585-08	GBF09	8	1.11	8.42	9.53	9.37	98.1	
N2585-09	GBF10	9	1.1	8.62	9.72	9.59	98.5	
N2585-10	GBF11	10	1.1	8.77	9.87	9.7	98.1	
N2585-11	GBF11D	11	1.1	8.77	9.87	9.7	98.1	
N2585-12	GBF11S	12	1.1	8.77	9.87	9.7	98.1	
N2585-13	GBF12	13	1.1	8.41	9.51	9.3	97.5	
N2585-14	GBF13	14	1.1	8.47	9.57	9.36	97.5	
N2585-15	GBF14	15	1.11	8.57	9.68	9.52	98.1	
N2585-16	GBF15	16	1.12	8.46	9.58	9.24	96.0	
N2585-17	GBF16	17	1.1	8.42	9.52	9.39	98.5	
N2585-18	GBF17	18	1.1	8.49	9.59	9.49	98.8	
N2585-19	GBF18	19	1.11	8.47	9.58	9.32	96.9	
N2585-20	GBF19	20	1.12	8.72	9.84	9.69	98.3	
N2585-21	GBF20	21	1.1	8.52	9.62	9.6	99.8	
N2585-22	GBF21	22	1.11	8.72	9.83	9.8	99.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$