



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB119879

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	04/28/2022	10:10
2	N2575-12		1	9	28.00	82.4	82.1	04/28/2022	10:40
3	N2575-12DUP		1	9	30.00	86.0	85.7	04/28/2022	11:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB119879

WorkList Name : fp42722-1 WorkList ID : 159069 Department : Wet-Chemistry Date : 04-27-2022 15:15:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/03/2022	Solid	N2575-12	Flash Point	Cool 4 deg C	PSEG03	K11	30899	04/26/2022	1010B

Date/Time 04/28/22 09:40
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: RJ CEXT-1a6)

Date/Time 04/28/22 11:15
 Raw Sample Received by: RJ CEXT-1a6)
 Raw Sample Relinquished by: KS(wc)