

LB119883


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53284



A= 0.0000e+000

B= 1.8883e-004

C= -3.4132e-002

Rho= 0.9999149

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.018	-0.018	88	0.000	88					-
0.2	0.200	0.219	0.019	1341	0.0 %	1341					9
2.5	2.500	2.562	0.062	13746	0.0 %	13746					2
5.0	5.000	4.907	-0.093	26166	0.0 %	26166					-2
7.5	7.500	7.503	0.003	39912	0.0 %	39912					0
10.0	10.000	10.028	0.028	53284	0.0 %	53284					0

LB119883 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Apr 2022 13:43:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	88	-	0.0000	7471B	PPB	28 Apr 2022 13:45:54	S	Std
0.2 - 1	50.2	1341	-	0.2000	7471B	PPB	28 Apr 2022 13:48:11	S	Std
2.5 - 1	52.5	13746	-	2.5000	7471B	PPB	28 Apr 2022 13:50:29	S	Std
5.0 - 1	55	26166	-	5.0000	7471B	PPB	28 Apr 2022 13:52:45	S	Std
7.5 - 1	57.5	39912	-	7.5000	7471B	PPB	28 Apr 2022 13:55:02	S	Std
10.0 - 1	510	53284	-	10.0000	7471B	PPB	28 Apr 2022 13:57:18	S	Std
ICV99 - 1	ICV99	21098	3.9498	-	7471B	PPB	28 Apr 2022 14:00:13	U	SMPL
ICB99 - 1	ICB99	-25	-0.0389	-	7471B	PPB	28 Apr 2022 14:02:28	U	SMPL
CCV49 - 1	CCV49	26474	4.9650	-	7471B	PPB	28 Apr 2022 14:04:42	U	SMPL
CCB49 - 1	CCB49	-34	-0.0406	-	7471B	PPB	28 Apr 2022 14:06:58	U	SMPL
CRA - 1	CRA	1319	0.2149	-	7471B	PPB	28 Apr 2022 14:09:14	U	SMPL
HighStd - 1	HighStd	53108	9.9943	-	7471B	PPB	28 Apr 2022 14:11:29	U	SMPL
ChkStd - 1	ChkStd	37726	7.0897	-	7471B	PPB	28 Apr 2022 14:13:44	U	SMPL
PB144453BL - 1	PBS	-46	-0.0428	-	7471B	PPB	28 Apr 2022 14:16:01	U	SMPL
PB144453BS - 1	LCSS	22604	4.2342	-	7471B	PPB	28 Apr 2022 14:18:18	U	SMPL
N2560-01 - 1	2827	52262	9.8346	-	7471B	PPB	28 Apr 2022 14:20:34	U	SMPL
N2564-01 - 1	PSC-164718	12868	2.3957	-	7471B	PPB	28 Apr 2022 14:22:50	U	SMPL
N2564-03 - 1	PSC-124019	7014	1.2903	-	7471B	PPB	28 Apr 2022 14:25:08	U	SMPL
N2569-01 - 1	OR-02-042622	3736	0.6713	-	7471B	PPB	28 Apr 2022 14:27:25	U	SMPL
N2572-01 - 1	MH-14S	965	0.1481	-	7471B	PPB	28 Apr 2022 14:29:39	U	SMPL
CCV50 - 1	CCV50	26400	4.9510	-	7471B	PPB	28 Apr 2022 14:31:54	U	SMPL
CCB50 - 1	CCB50	1	-0.0339	-	7471B	PPB	28 Apr 2022 14:34:10	U	SMPL
N2572-04 - 1	MH-14D	642	0.0871	-	7471B	PPB	28 Apr 2022 14:36:26	U	SMPL
N2588-01 - 1	MH-1	9733	1.8038	-	7471B	PPB	28 Apr 2022 14:38:41	U	SMPL
N2588-03 - 1	MH-2	804	0.1177	-	7471B	PPB	28 Apr 2022 14:40:57	U	SMPL
N2598-01 - 1	OS-DRUM-COMP	3322	0.5932	-	7471B	PPB	28 Apr 2022 14:43:12	U	SMPL
N2599-01 - 1	M-1	4354	0.7880	-	7471B	PPB	28 Apr 2022 14:45:28	U	SMPL
N2600-01 - 1	MH-15S	509	0.0620	-	7471B	PPB	28 Apr 2022 14:47:44	U	SMPL
N2600-04 - 1	MH-15D	222	0.0078	-	7471B	PPB	28 Apr 2022 14:50:01	U	SMPL
N2600-04DUP - 1	MH-15DDUP	215	0.0065	-	7471B	PPB	28 Apr 2022 14:52:18	U	SMPL
N2600-04MS - 1	MH-15DMS	20977	3.9270	-	7471B	PPB	28 Apr 2022 14:54:34	U	SMPL
N2600-04MSD - 1	MH-15DMSD	20685	3.8718	-	7471B	PPB	28 Apr 2022 14:56:50	U	SMPL
CCV51 - 1	CCV51	25687	4.8164	-	7471B	PPB	28 Apr 2022 14:59:05	U	SMPL
CCB51 - 1	CCB51	-13	-0.0366	-	7471B	PPB	28 Apr 2022 15:01:21	U	SMPL
N2600-04LX5 - 1		94	-0.0164	-	7471B	PPB	28 Apr 2022 15:03:37	U	SMPL
N2600-04A - 1		22506	4.2157	-	7471B	PPB	28 Apr 2022 15:05:52	U	SMPL
CCV52 - 1	CCV52	26152	4.9042	-	7471B	PPB	28 Apr 2022 15:08:07	U	SMPL
CCB52 - 1	CCB52	-32	-0.0402	-	7471B	PPB	28 Apr 2022 15:10:23	U	SMPL