



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/29/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 04/28/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 04/29/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119885

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2588-05	MH-1-EPH	1	1.19	8.58	9.77	8.83	89.0	
N2588-06	MH-2-EPH	2	1.14	8.61	9.75	8.89	90.0	
N2602-01	12-15-SB1-BK-PRISON	3	1.14	8.48	9.62	8.65	88.6	
N2602-02	12-15-SB5-BK-PRISON	4	0.94	8.78	9.72	8.53	86.4	
N2602-03	12-15-SB4-BK-PRISON	5	0.9	8.95	9.85	8.76	87.8	
N2602-04	12-20-SB2-BK-PRISON	6	0.94	8.90	9.84	8.94	89.9	
N2602-05	10-16-SB3-BK-PRISON	7	0.92	8.70	9.62	9.25	95.7	
N2604-01	NB-07-042822	16	1.14	8.40	9.54	8.3	85.2	
N2604-02	NB-07-042822-EPH-2	17	1.14	8.74	9.88	8.22	81.0	
N2605-01	TR-04-042822	18	0.94	8.61	9.55	8.73	90.5	
N2605-02	TR-04-042822-EPH-2	19	0.94	8.82	9.76	8.68	87.8	
N2605-03	TR-05-042822	20	0.94	8.97	9.91	9.74	98.1	
N2605-04	TR-05-042822-EPH-2	21	0.91	8.93	9.84	9.66	98.0	
N2605-05	TR-06-042822	22	1.19	8.56	9.75	9.05	91.8	
N2605-06	TR-06-042822-EPH-2	23	0.94	8.83	9.77	9.22	93.8	
N2606-01	WC-3	38	0.94	8.61	9.55	8.88	92.2	
N2607-01	HR-02-042822	24	1.17	8.53	9.7	9.35	95.9	
N2607-02	HR-02-042822-EPH-2	25	0.91	9.01	9.92	9.77	98.3	
N2609-01	WC-1	8	0.92	8.98	9.9	8.6	85.5	
N2609-03	WC-1	9	0.92	8.98	9.9	8.6	85.5	
N2609-04	WC-2	10	0.93	9.04	9.97	8.89	88.1	
N2609-06	WC-2	11	0.93	9.04	9.97	8.89	88.1	
N2609-07	WC-3	12	0.91	9.07	9.98	8.55	84.2	
N2609-09	WC-3	13	0.91	9.07	9.98	8.55	84.2	
N2609-10	WC-4	14	1.14	8.77	9.91	8.32	81.9	
N2609-12	WC-4	15	1.14	8.77	9.91	8.32	81.9	
N2610-01	HR-03-042822	26	1.14	8.49	9.63	8.93	91.8	
N2610-02	HR-03-042822-EPH-2	27	1.14	8.49	9.63	8.93	91.8	



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/29/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 04/28/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 04/29/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119885

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2611-01	JC-01-042822	28	1.18	8.50	9.68	9.13	93.5	
N2611-02	JC-01-042822-EPH-2	29	1.18	8.50	9.68	9.13	93.5	
N2616-01	MH-18S	30	1.18	8.60	9.78	7.54	74.0	
N2616-03	MH-18S-EPH	31	1.19	8.73	9.92	7.19	68.7	
N2616-04	MH-18D	32	0.93	8.93	9.86	7.24	70.7	
N2616-06	MH-18D-EPH	33	0.94	8.68	9.62	7.05	70.4	
N2616-07	MH-20S	34	0.93	8.96	9.89	9.25	92.9	
N2616-09	MH-20S-EPH	35	0.93	8.63	9.56	8.8	91.2	
N2616-10	MH-20D	36	0.91	9.07	9.98	9.52	94.9	
N2616-12	MH-20D-EPH	37	0.94	8.89	9.83	9.29	93.9	
N2618-01	SB-01	39	0.92	8.92	9.84	9.09	91.6	
N2618-02	SB-02	40	0.92	8.75	9.67	8.93	91.5	
N2618-03	SB-03	41	0.91	8.59	9.5	8.6	89.5	
N2618-04	SB-04	42	0.97	8.75	9.72	8.57	86.9	
N2618-05	SB-05	43	0.94	9.03	9.97	8.00	78.2	
N2618-06	SB-06	44	0.92	8.80	9.72	7.86	78.9	
N2618-07	SB-07	45	0.92	8.96	9.88	8.54	85.0	
N2618-08	SB-08	46	1.11	8.86	9.97	7.88	76.4	
N2618-09	SB-09	47	0.92	8.58	9.5	7.21	73.3	
N2618-09D	SB-09DUP	48	0.93	8.60	9.53	7.23	73.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$