



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119890

Slope : 99.2

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	04/28/2022	17:10
2	CAL2	1	Water	NA	NA	20.6	7.00	04/28/2022	17:11
3	CAL3	1	Water	NA	NA	20.7	10.03	04/28/2022	17:12
4	ICV	1	Water	NA	NA	20.7	6.99	04/28/2022	17:15
5	CCV1	1	Water	NA	NA	20.7	2.01	04/28/2022	17:16
6	N2600-02	1	Solid	20.02	20	21.4	9.21	04/28/2022	17:19
7	N2600-05	1	Solid	20.03	20	20.9	8.48	04/28/2022	17:20
8	N2606-02	1	Solid	20.01	20	21.8	7.26	04/28/2022	17:22
9	N2609-02	1	Solid	20.03	20	21.0	8.76	04/28/2022	17:25
10	N2609-05	1	Solid	20.04	20	21.3	7.32	04/28/2022	17:26
11	N2609-08	1	Solid	20.01	20	21.5	7.92	04/28/2022	17:30
12	N2609-11	1	Solid	20.02	20	21.2	9.26	04/28/2022	17:32
13	N2616-02	1	Solid	20.04	20	21.7	7.75	04/28/2022	17:35
14	N2616-05	1	Solid	20.02	20	21.5	6.99	04/28/2022	17:36
15	N2616-08	1	Solid	20.03	20	20.8	6.51	04/28/2022	17:39
16	CCV2	1	Water	NA	NA	20.2	12.02	04/28/2022	17:40
17	N2616-11	1	Solid	20.04	20	20.8	5.85	04/28/2022	17:42
18	N2616-11DUP	1	Solid	20.03	20	20.9	5.86	04/28/2022	17:43
19	CCV3	1	Water	NA	NA	20.7	2.01	04/28/2022	17:45

28119890

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2600 WorkList ID : 159104 Department : Wet-Chemistry Date : 04-28-2022 13:07:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
04/27/2022	Solid	N2600-02	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-15S	04/27/2022	9045D
04/27/2022	Solid	N2600-05	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-15D	04/27/2022	9045D
04/28/2022	Solid	N2606-02	Corrosivity	Cool 4 deg C	DVIR01	K11	WC-3	04/28/2022	9045D
04/28/2022	Solid	N2609-02	Corrosivity	Cool 4 deg C	PSEG03	K11	WC-1	04/28/2022	9045D
04/28/2022	Solid	N2609-05	Corrosivity	Cool 4 deg C	PSEG03	K11	WC-2	04/28/2022	9045D
04/28/2022	Solid	N2609-08	Corrosivity	Cool 4 deg C	PSEG03	K11	WC-3	04/28/2022	9045D
04/28/2022	Solid	N2609-11	Corrosivity	Cool 4 deg C	PSEG03	K11	WC-4	04/28/2022	9045D
04/28/2022	Solid	N2616-02	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-18S	04/28/2022	9045D
04/28/2022	Solid	N2616-05	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-18D	04/28/2022	9045D
04/28/2022	Solid	N2616-08	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-20S	04/28/2022	9045D
04/28/2022	Solid	N2616-11	Corrosivity	Cool 4 deg C	PSEG03	K11	MH-20D	04/28/2022	9045D

Date/Time 04/28/22 13:40
 Raw Sample Received by: 10 (GNC)
 Raw Sample Relinquished by: JD CSM

Date/Time 04/28/22 17:40
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: 10 (GNC)