



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 5/2/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 04/29/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:21
Out Date: 04/30/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119896

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2562-01	EW461	1	0.91	9.01	9.92	8.57	85.0	
N2562-02	EW462	2	0.9	9.06	9.96	9.25	92.2	
N2562-03	EW463	3	0.92	8.98	9.9	8.05	79.4	
N2562-04	EW464	4	0.92	9.05	9.97	8.66	85.5	
N2562-05	EW466	5	0.9	9.08	9.98	8.15	79.8	
N2562-06	EW467	6	0.92	9.05	9.97	9.11	90.5	
N2562-07	EW468	7	0.94	8.98	9.92	9.00	89.8	
N2562-08	EW469	8	0.92	9.02	9.94	9.00	89.6	
N2562-09	VHBLK001	9	1.00	1.00	2.00	2.00	100.0	VHBLK
N2562-10	EW470	10	0.93	8.62	9.55	8.32	85.7	
N2562-11	EW472	11	0.92	8.89	9.81	8.95	90.3	
N2562-12	EW473	12	0.94	8.74	9.68	7.78	78.3	
N2562-13	EW474	13	0.9	9.08	9.98	9.52	94.9	
N2562-14	EW475	14	0.93	8.71	9.64	7.67	77.4	
N2562-15	EW476	15	0.95	9.00	9.95	9.18	91.4	
N2562-16	EW477	16	0.91	8.74	9.65	8.00	81.1	
N2562-17	EW478	17	0.93	8.90	9.83	8.12	80.8	
N2562-19	EW478MS	18	0.93	8.90	9.83	8.12	80.8	
N2562-20	EW478MSD	19	0.93	8.90	9.83	8.12	80.8	
N2562-21	EW480	20	0.91	8.77	9.68	8.75	89.4	
N2562-22	EW481	21	0.93	9.05	9.98	8.97	88.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$