



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119900

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	04/29/2022	12:35
2	CAL2	1	Water	NA	NA	20.4	7.00	04/29/2022	12:36
3	CAL3	1	Water	NA	NA	20.5	10.03	04/29/2022	12:37
4	ICV	1	Water	NA	NA	20.5	6.99	04/29/2022	12:39
5	CCV1	1	Water	NA	NA	20.6	2.01	04/29/2022	12:40
6	N2598-03	1	Water	NA	NA	20.4	8.06	04/29/2022	12:42
7	N2624-01	1	Water	NA	NA	20.1	7.34	04/29/2022	12:45
8	N2624-01DUP	1	Water	NA	NA	20.2	7.35	04/29/2022	12:46
9	CCV2	1	Water	NA	NA	20.6	12.03	04/29/2022	12:48

VB 119900

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2624 WorkList ID : 159144 Department : Wet-Chemistry Date : 04-29-2022 12:26:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/27/2022	Water	N2598-03	pH	Cool 4 deg C	PSEG03	L11	OW-TOTE-COMP	04/27/2022	9040C
04/29/2022	Water	N2624-01	pH	Cool 4 deg C	PSEG03		N47992	04/29/2022	9040C

Date/Time 04/29/22 12:30
Raw Sample Received by: JN(wf)
Raw Sample Relinquished by: JR(sm)

Date/Time 04/29/22 14:00
Raw Sample Received by: JR(sm)
Raw Sample Relinquished by: JR(wc)