



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119917

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/29/2022	17:00
2	CAL2	1	Water	NA	NA	20.3	7.00	04/29/2022	17:01
3	CAL3	1	Water	NA	NA	20.4	10.03	04/29/2022	17:02
4	ICV	1	Water	NA	NA	20.4	6.98	04/29/2022	17:05
5	CCV1	1	Water	NA	NA	20.2	2.02	04/29/2022	17:20
6	N2642-01	1	Water	NA	NA	21.5	8.65	04/29/2022	17:22
7	N2642-02	1	Water	NA	NA	21.6	8.69	04/29/2022	17:24
8	N2643-01	1	Water	NA	NA	21.6	7.56	04/29/2022	17:26
9	N2643-01DUP	1	Water	NA	NA	21.7	7.57	04/29/2022	17:27
10	CCV2	1	Water	NA	NA	20.6	12.03	04/29/2022	17:30

WORKLIST(Hardcopy Internal Chain)

VB 119914

WorkList Name : ph n2642 WorkList ID : 159163 Department : Wet-Chemistry Date : 04-29-2022 17:11:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
04/29/2022	Water	N2642-01	pH	Cool 4 deg C	PSEG03	L11	30988	04/29/2022	9040C
04/29/2022	Water	N2642-02	pH	Cool 4 deg C	PSEG03	L11	30989	04/29/2022	9040C
04/29/2022	Water	N2643-01	pH	Cool 4 deg C	PSEG03	L11	31187	04/29/2022	9040C

Date/Time 04/29/22 17:20
 Raw Sample Received by: JS(wc)
 Raw Sample Relinquished by: JS(wc)

Date/Time 04/29/22 18:20
 Raw Sample Received by: JS(wc)
 Raw Sample Relinquished by: JS(wc)