


LB119932

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3573e-004

C= -5.8783e-002

Rho= 0.9997446

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.044	-0.044	63	0.000	63					-16
0.2	0.200	0.232	0.032	1234	0.0 %	1234					-2
2.5	2.500	2.439	-0.061	10597	0.0 %	10597					-3
5.0	5.000	5.167	0.167	22168	0.0 %	22168					-1
7.5	7.500	7.415	-0.085	31707	0.0 %	31707					-1
10.0	10.000	9.990	-0.010	42631	0.0 %	42631					0

LB119932 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 May 2022 10:45:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	63	-	0.0000	7470A	PPB	02 May 2022 10:48:35	S	Std
0.2 - 1	50.2	1234	-	0.2000	7470A	PPB	02 May 2022 10:50:52	S	Std
2.5 - 1	52.5	10597	-	2.5000	7470A	PPB	02 May 2022 10:53:09	S	Std
5.0 - 1	55	22168	-	5.0000	7470A	PPB	02 May 2022 10:55:26	S	Std
7.5 - 1	57.5	31707	-	7.5000	7470A	PPB	02 May 2022 10:57:42	S	Std
10.0 - 1	510	42631	-	10.0000	7470A	PPB	02 May 2022 10:59:58	S	Std
ICV04 - 1	ICV04	15988	3.7100	-	7470A	PPB	02 May 2022 11:02:44	U	SMPL
ICB04 - 1	ICB04	-1	-0.0590	-	7470A	PPB	02 May 2022 11:04:58	U	SMPL
CCV62 - 1	CCV62	20963	4.8827	-	7470A	PPB	02 May 2022 11:09:55	U	SMPL
CCB62 - 1	CCB62	19	-0.0543	-	7470A	PPB	02 May 2022 11:12:09	U	SMPL
CRA - 1	CRA	1224	0.2297	-	7470A	PPB	02 May 2022 11:14:24	U	SMPL
HighStd - 1	HighStd	42121	9.8702	-	7470A	PPB	02 May 2022 11:16:39	U	SMPL
ChkStd - 1	ChkStd	29627	6.9251	-	7470A	PPB	02 May 2022 11:18:54	U	SMPL
PB144476BL - 1	PBW	-93	-0.0807	-	7470A	PPB	02 May 2022 11:21:12	U	SMPL
PB144476BS - 1	LCSW	15964	3.7043	-	7470A	PPB	02 May 2022 11:26:03	U	SMPL
N2588-02 - 1	MH-1	-40	-0.0682	-	7470A	PPB	02 May 2022 11:28:18	U	SMPL
N2588-04 - 1	MH-2	3	-0.0581	-	7470A	PPB	02 May 2022 11:30:34	U	SMPL
N2600-02 - 1	MH-15S	-148	-0.0937	-	7470A	PPB	02 May 2022 11:32:51	U	SMPL
N2600-05 - 1	MH-15D	54	-0.0461	-	7470A	PPB	02 May 2022 11:35:08	U	SMPL
N2609-02 - 1	WC-1	39	-0.0496	-	7470A	PPB	02 May 2022 11:37:23	U	SMPL
CCV63 - 1	CCV63	21746	5.0673	-	7470A	PPB	02 May 2022 11:39:38	U	SMPL
CCB63 - 1	CCB63	13	-0.0557	-	7470A	PPB	02 May 2022 11:41:54	U	SMPL
N2609-05 - 1	WC-2	41	-0.0491	-	7470A	PPB	02 May 2022 11:44:09	U	SMPL
N2609-08 - 1	WC-3	41	-0.0491	-	7470A	PPB	02 May 2022 11:46:25	U	SMPL
N2609-11 - 1	WC-4	43	-0.0486	-	7470A	PPB	02 May 2022 11:48:40	U	SMPL
N2616-02 - 1	MH-18S	68	-0.0428	-	7470A	PPB	02 May 2022 11:50:55	U	SMPL
N2616-05 - 1	MH-18D	42	-0.0489	-	7470A	PPB	02 May 2022 11:53:11	U	SMPL
N2616-08 - 1	MH-20S	58	-0.0451	-	7470A	PPB	02 May 2022 11:55:28	U	SMPL
N2616-11 - 1	MH-20D	61	-0.0444	-	7470A	PPB	02 May 2022 12:02:14	U	SMPL
N2616-11DUP - 1	MH-20DDUP	56	-0.0456	-	7470A	PPB	02 May 2022 12:04:31	U	SMPL
N2616-11MS - 1	MH-20DMS	14447	3.3468	-	7470A	PPB	02 May 2022 12:06:47	U	SMPL
N2616-11MSD - 1	MH-20DMSD	14881	3.4491	-	7470A	PPB	02 May 2022 12:09:02	U	SMPL
CCV64 - 1	CCV64	21983	5.1232	-	7470A	PPB	02 May 2022 12:11:18	U	SMPL
CCB64 - 1	CCB64	45	-0.0482	-	7470A	PPB	02 May 2022 12:13:33	U	SMPL
PB144451TB - 1	PB144451TB	23	-0.0534	-	7470A	PPB	02 May 2022 12:15:48	U	SMPL
LB119932LX5 - 1		122	-0.0300	-	7470A	PPB	02 May 2022 12:18:03	U	SMPL
LB119932A - 1		15951	3.7013	-	7470A	PPB	02 May 2022 12:20:19	U	SMPL
CCV65 - 1	CCV65	21584	5.0291	-	7470A	PPB	02 May 2022 12:22:34	U	SMPL
CCB65 - 1	CCB65	27	-0.0524	-	7470A	PPB	02 May 2022 12:24:49	U	SMPL