

7471B


LB119939

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55122



A= 0.0000e+000

B= 1.8210e-004

C= -7.0240e-002

Rho= 0.9998934

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/b)
0.0	0.000	-0.071	-0.071	-2	0.000	-2					-
0.2	0.200	0.228	0.028	1637	0.0 %	1637					14
2.5	2.500	2.503	0.003	14130	0.0 %	14130					0
5.0	5.000	5.099	0.099	28388	0.0 %	28388					2
7.5	7.500	7.473	-0.027	41421	0.0 %	41421					0
10.0	10.000	9.968	-0.032	55122	0.0 %	55122					0

LB119939 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 02 May 2022 13:38:37					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.0000	7471B	PPB	02 May 2022 13:43:57	S	Std
0.2 - 1	50.2	1637	-	0.2000	7471B	PPB	02 May 2022 13:46:14	S	Std
2.5 - 1	52.45	14130	-	2.5000	7471B	PPB	02 May 2022 13:48:31	S	Std
5.0 - 1	55	28388	-	5.0000	7471B	PPB	02 May 2022 13:50:49	S	Std
7.5 - 1	57.5	41421	-	7.5000	7471B	PPB	02 May 2022 13:53:05	S	Std
10.0 - 1	510	55122	-	10.0000	7471B	PPB	02 May 2022 13:58:00	S	Std
ICV05 - 1	ICV05	22088	3.9521	-	7471B	PPB	02 May 2022 14:00:47	U	SMPL
ICB05 - 1	ICB05	-91	-0.0868	-	7471B	PPB	02 May 2022 14:03:01	U	SMPL
CCV66 - 1	CCV66	28353	5.0930	-	7471B	PPB	02 May 2022 14:05:16	U	SMPL
CCB66 - 1	CCB66	-100	-0.0885	-	7471B	PPB	02 May 2022 14:07:32	U	SMPL
CRA - 1	CRA	1623	0.2253	-	7471B	PPB	02 May 2022 14:09:47	U	SMPL
HighStd - 1	HighStd	55061	9.9566	-	7471B	PPB	02 May 2022 14:12:02	U	SMPL
ChkStd - 1	ChkStd	38573	6.9541	-	7471B	PPB	02 May 2022 14:14:17	U	SMPL
PB144522BL - 1	PBS	-126	-0.0932	-	7471B	PPB	02 May 2022 14:16:34	U	SMPL
PB144522BS - 1	LCSS	19489	3.4788	-	7471B	PPB	02 May 2022 14:18:51	U	SMPL
N2619-23 - 1	SS11-13-FT	2602	0.4036	-	7471B	PPB	02 May 2022 14:21:07	U	SMPL
N2641-01 - 1	TP-21	1361	0.1776	-	7471B	PPB	02 May 2022 14:23:23	U	SMPL
N2641-01DUP - 1	TP-21DUP	1287	0.1641	-	7471B	PPB	02 May 2022 14:25:40	U	SMPL
N2641-01MS - 1	TP-21MS	20418	3.6480	-	7471B	PPB	02 May 2022 14:27:57	U	SMPL
N2641-01MSD - 1	TP-21MSD	21399	3.8266	-	7471B	PPB	02 May 2022 14:30:12	U	SMPL
CCV67 - 1	CCV67	28288	5.0811	-	7471B	PPB	02 May 2022 14:32:28	U	SMPL
CCB67 - 1	CCB67	-85	-0.0857	-	7471B	PPB	02 May 2022 14:34:43	U	SMPL
N2641-04 - 1	TP-24	1496	0.2022	-	7471B	PPB	02 May 2022 14:36:59	U	SMPL
N2643-03 - 1	31487	329	-0.0103	-	7471B	PPB	02 May 2022 14:39:14	U	SMPL
N2643-04 - 1	2822A	931	0.0993	-	7471B	PPB	02 May 2022 14:41:29	U	SMPL
PB144523BL - 1	PBS	13	-0.0679	-	7471B	PPB	02 May 2022 14:43:44	U	SMPL
PB144523BS - 1	LCSS	19511	3.4828	-	7471B	PPB	02 May 2022 14:46:00	U	SMPL
N2619-01 - 1	SS01-1-2-FT	1116	0.1330	-	7471B	PPB	02 May 2022 14:48:17	U	SMPL
N2619-02 - 1	SB01-13-14-FT	213	-0.0315	-	7471B	PPB	02 May 2022 14:50:33	U	SMPL
N2619-03 - 1	SS02-1-2-FT	6888	1.1841	-	7471B	PPB	02 May 2022 14:52:50	U	SMPL
N2619-04 - 1	SB02-13-14-FT	236	-0.0273	-	7471B	PPB	02 May 2022 14:55:07	U	SMPL
N2619-05 - 1	SS03-1-2-FT	4798	0.8035	-	7471B	PPB	02 May 2022 14:57:22	U	SMPL
CCV68 - 1	CCV68	28371	5.0963	-	7471B	PPB	02 May 2022 14:59:38	U	SMPL
CCB68 - 1	CCB68	-84	-0.0855	-	7471B	PPB	02 May 2022 15:01:53	U	SMPL
N2619-06 - 1	SB03-13-14-FT	587	0.0367	-	7471B	PPB	02 May 2022 15:04:08	U	SMPL
N2619-07 - 1	SS04-1.5-2.5-FT	2234	0.3366	-	7471B	PPB	02 May 2022 15:06:24	U	SMPL
N2619-08 - 1	SB04-14-15-FT	3988	0.6560	-	7471B	PPB	02 May 2022 15:08:39	U	SMPL
N2619-09 - 1	SS05-1-2-FT	1980	0.2903	-	7471B	PPB	02 May 2022 15:10:54	U	SMPL
N2619-10 - 1	SB05-13-15-FT	2948	0.4666	-	7471B	PPB	02 May 2022 15:13:09	U	SMPL
N2619-10DUP - 1	SB05-13-15-FTDUP	3233	0.5185	-	7471B	PPB	02 May 2022 15:15:26	U	SMPL
N2619-11 - 1	N2619-10MS	22446	4.0173	-	7471B	PPB	02 May 2022 15:17:42	U	SMPL
N2619-12 - 1	N2619-10MSD	22434	4.0151	-	7471B	PPB	02 May 2022 15:19:59	U	SMPL
N2619-13 - 1	SS06-1-2-FT	989	0.1099	-	7471B	PPB	02 May 2022 15:22:16	U	SMPL
N2619-14 - 1	SB06-12-14-FT	146	-0.0437	-	7471B	PPB	02 May 2022 15:24:32	U	SMPL
CCV69 - 1	CCV69	28713	5.1585	-	7471B	PPB	02 May 2022 15:26:47	U	SMPL
CCB69 - 1	CCB69	-57	-0.0806	-	7471B	PPB	02 May 2022 15:29:02	U	SMPL
N2619-15 - 1	SS07-1-2-FT	3435	0.5553	-	7471B	PPB	02 May 2022 15:31:18	U	SMPL
N2619-16 - 1	SB07-13-14-FT	292	-0.0171	-	7471B	PPB	02 May 2022 15:33:33	U	SMPL
N2619-17 - 1	SS08-1-2-FT	7176	1.2365	-	7471B	PPB	02 May 2022 15:35:48	U	SMPL
N2619-18 - 1	SB08-13-14-FT	517	0.0239	-	7471B	PPB	02 May 2022 15:38:04	U	SMPL
N2619-19 - 1	SS09-1-2-FT	16937	3.0141	-	7471B	PPB	02 May 2022 15:40:19	U	SMPL
N2619-20 - 1	SB09-13-14-FT	143	-0.0442	-	7471B	PPB	02 May 2022 15:42:35	U	SMPL
N2619-21 - 1	SS10-1-2-FT	1889	0.2738	-	7471B	PPB	02 May 2022 15:44:51	U	SMPL
N2619-22 - 1	SB10-13-14-FT	208	-0.0324	-	7471B	PPB	02 May 2022 15:47:08	U	SMPL
CCV70 - 1	CCV70	29024	5.2152	-	7471B	PPB	02 May 2022 15:49:25	U	SMPL
CCB70 - 1	CCB70	-65	-0.0821	-	7471B	PPB	02 May 2022 15:51:40	U	SMPL
N2641-01LX5 - 1		211	-0.0318	-	7471B	PPB	02 May 2022 15:53:56	U	SMPL
N2641-01A - 1		22740	4.0708	-	7471B	PPB	02 May 2022 15:56:12	U	SMPL
N2619-10LX5 - 1		502	0.0212	-	7471B	PPB	02 May 2022 15:58:27	U	SMPL
N2619-10A - 1		23853	4.2735	-	7471B	PPB	02 May 2022 16:00:42	U	SMPL
CCV71 - 1	CCV71	28402	5.1019	-	7471B	PPB	02 May 2022 16:02:57	U	SMPL
CCB71 - 1	CCB71	-71	-0.0832	-	7471B	PPB	02 May 2022 16:05:14	U	SMPL