



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 5/5/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 05/02/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 05/03/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119942

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2640-01	IDW-S-10-042922	1	0.9	8.95	9.85	6.85	66.5	
N2647-01	357	60	0.93	9.03	9.96	5.66	52.4	GEL MATRIX
N2647-02	357-EPH-2	61	0.94	8.77	9.71	5.22	48.8	GEL MATRIX
N2648-01	RBR-346	62	0.91	8.95	9.86	4.9	44.6	GEL MATRIX
N2648-02	RBR-346-EPH-2	63	0.95	8.87	9.82	3.33	26.8	GEL MATRIX
N2649-01	WC-1	72	0.92	8.80	9.72	8.8	89.5	
N2650-01	ETGI-349	2	0.93	9.04	9.97	7.41	71.7	GEL MATRIX
N2650-02	ETGI-349-EPH-2	3	0.93	9.04	9.97	7.41	71.7	GEL MATRIX
N2650-03	ETGI-362	4	0.91	9.02	9.93	7.48	72.8	
N2650-04	ETGI-362-EPH-2	5	0.91	9.02	9.93	7.48	72.8	
N2651-01	CL-54-05-02-22-A	6	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2651-02	CL-54-05-02-22-B	7	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2651-03	CL-55-05-02-22-A	8	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2651-04	CL-55-05-02-22-B	9	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2651-05	CL-56-05-02-22-A	10	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2651-06	CL-56-05-02-22-B	11	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2651-07	CL-57-05-02-22-A	12	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2651-08	CL-57-05-02-22-B	13	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2652-05	SVOC-GPC-BLANK	14	1.00	1.00	2.00	2.00	100.0	
N2652-06	PEST-GPC-BLANK	15	1.00	1.00	2.00	2.00	100.0	
N2652-07	PEST-GPC-BLANK-SPIKE	16	1.00	1.00	2.00	2.00	100.0	
N2652-08	PCB-GPC-BLANK	17	1.00	1.00	2.00	2.00	100.0	
N2652-09	PCB-GPC-BLANK-SPIKE	18	1.00	1.00	2.00	2.00	100.0	
N2652-10	SVOC-GPC2-BLANK	19	1.00	1.00	2.00	2.00	100.0	
N2652-11	PEST-GPC2-BLANK	20	1.00	1.00	2.00	2.00	100.0	
N2652-12	PEST-GPC2-BLANK-SPIKE	21	1.00	1.00	2.00	2.00	100.0	
N2652-13	PCB-GCP2-BLANK	22	1.00	1.00	2.00	2.00	100.0	
N2652-14	PCB-GCP2-BLANK-SPIKE	23	1.00	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 5/5/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 05/02/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 05/03/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119942

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2653-01	31591	64	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N2653-02	31592	65	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N2653-03	31593	66	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N2653-04	31594	67	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N2653-05	31595	68	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N2653-06	31596	69	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N2653-07	31597	70	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N2654-01	ROOF-FLASHING-CAULK	24	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N2654-02	ROOF-SOFFIT-CAULK	25	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N2654-03	DUCT-PUTTY	26	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N2655-01	CL-01-050222	27	0.92	9.06	9.98	9.02	89.4	
N2655-02	CL-01-050222-EPH-2	28	0.92	9.06	9.98	9.02	89.4	
N2656-01	PT-AN-SOIL	29	1.00	1.00	2.00	2.00	100.0	
N2656-02	PT-CORR-SOIL	30	1.00	1.00	2.00	2.00	100.0	
N2656-03	PT-CORR-SOIL	31	1.00	1.00	2.00	2.00	100.0	
N2656-04	PT-CN-SOIL	32	1.00	1.00	2.00	2.00	100.0	
N2656-05	PT-CN-SOIL	33	1.00	1.00	2.00	2.00	100.0	
N2656-06	PT-FP-SOIL	34	1.00	1.00	2.00	2.00	100.0	
N2656-07	PT-CR6-SOIL	35	1.00	1.00	2.00	2.00	100.0	
N2656-08	PT-NUT-SOIL	36	1.00	1.00	2.00	2.00	100.0	
N2656-09	PT-NUT-SOIL	37	1.00	1.00	2.00	2.00	100.0	
N2656-10	PT-OGR-SOIL	38	1.00	1.00	2.00	2.00	100.0	
N2656-11	PT-MET-SOIL	39	1.00	1.00	2.00	2.00	100.0	
N2656-12	PT-BNA-SOIL	40	1.00	1.00	2.00	2.00	100.0	
N2656-13	PT-TRIAZINE-SOIL	41	1.00	1.00	2.00	2.00	100.0	
N2656-14	PT-PAH-SOIL	42	1.00	1.00	2.00	2.00	100.0	
N2656-15	PT-DIES-SOIL	43	1.00	1.00	2.00	2.00	100.0	
N2656-16	PT-GAS-SOIL	44	1.00	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 5/5/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 05/02/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 05/03/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119942

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2656-17	PT-NJEPH-SOIL	45	1.00	1.00	2.00	2.00	100.0	
N2656-18	PT-HERB-SOIL	46	1.00	1.00	2.00	2.00	100.0	
N2656-19	PT-PCB-SOIL	47	1.00	1.00	2.00	2.00	100.0	
N2656-20	PT-PCBO-SOIL	48	1.00	1.00	2.00	2.00	100.0	
N2656-21	PT-PEST-SOIL	49	1.00	1.00	2.00	2.00	100.0	
N2656-22	PT-CHLR-SOIL	50	1.00	1.00	2.00	2.00	100.0	
N2656-23	PT-TXP-SOIL	51	1.00	1.00	2.00	2.00	100.0	
N2656-24	PT-VOA-SOIL	52	1.00	1.00	2.00	2.00	100.0	
N2656-25	PT-SOL-SOIL	53	0.92	9.04	9.96	7.82	76.3	
N2657-01	HOTTEST-1	54	0.93	9.03	9.96	9.34	93.1	
N2657-02	HOTTEST-2	55	0.94	8.94	9.88	8.87	88.7	
N2659-01	22L097-A	56	0.91	8.63	9.54	8.47	87.6	
N2659-02	22L097-B	57	0.91	8.59	9.5	8.4	87.2	
N2659-03	22L097-C	58	0.94	8.96	9.9	8.71	86.7	
N2659-04	22L097-D	59	0.97	8.83	9.8	8.54	85.7	
N2663-01	TP-25	73	0.95	8.99	9.94	9.12	90.9	
N2663-03	TP-25-EPH	74	0.92	8.94	9.86	8.74	87.5	
N2663-03D	TP-25-EPHDUP	75	0.93	8.95	9.88	8.72	87.0	
N2664-01	S-6	71	1.19	8.50	9.69	7.71	76.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$