



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119946

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/02/2022	16:05
2	CAL2	1	Water	NA	NA	20.3	7.00	05/02/2022	16:06
3	CAL3	1	Water	NA	NA	20.3	10.03	05/02/2022	16:07
4	ICV	1	Water	NA	NA	20.4	6.99	05/02/2022	16:10
5	CCV1	1	Water	NA	NA	20.4	2.01	05/02/2022	16:12
6	N2659-01	1	Solid	20.02	20	21.7	11.84	05/02/2022	16:15
7	N2659-01DUP	1	Solid	20.03	20	21.8	11.85	05/02/2022	16:16
8	N2659-02	1	Solid	20.02	20	21.4	12.02	05/02/2022	16:20
9	N2659-03	1	Solid	20.03	20	21.5	11.88	05/02/2022	16:22
10	N2659-04	1	Solid	20.04	20	21.2	11.85	05/02/2022	16:25
11	CCV2	1	Water	NA	NA	20.6	12.02	05/02/2022	16:26

283 119946

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2659 WorkList ID : 159210 Department : Wet-Chemistry Date : 05-02-2022 15:55:14

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
05/02/2022	Solid	N2659-01	pH	Cool 4 deg C	ENV/81	K22	ZLLO97-A	05/02/2022 9045D
05/02/2022	Solid	N2659-02	pH	Cool 4 deg C	ENV/81	K22	ZLLO97-B	05/02/2022 9045D
05/02/2022	Solid	N2659-03	pH	Cool 4 deg C	ENV/81	K22	ZLLO97-C	05/02/2022 9045D
05/02/2022	Solid	N2659-04	pH	Cool 4 deg C	ENV/81	K22	ZLLO97-D	05/02/2022 9045D

Date/Time 05/02/22 16:00
Raw Sample Received by: JR (SM)
Raw Sample Relinquished by: JR (SM)

Date/Time 05/02/22 17:00
Raw Sample Received by: JR (SM)
Raw Sample Relinquished by: JR (WC)