



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119948

Slope : 98.8

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	05/02/2022	17:10
2	CAL2	1	Water	NA	NA	20.4	7.00	05/02/2022	17:11
3	CAL3	1	Water	NA	NA	20.4	10.03	05/02/2022	17:12
4	ICV	1	Water	NA	NA	20.5	6.99	05/02/2022	17:15
5	CCV1	1	Water	NA	NA	20.5	2.01	05/02/2022	17:20
6	N2641-03	1	Solid	20.03	20	21.0	7.45	05/02/2022	17:22
7	N2641-06	1	Solid	20.01	20	20.4	9.38	05/02/2022	17:25
8	N2644-03	1	Solid	20.01	20	21.0	6.38	05/02/2022	17:26
9	N2663-02	1	Solid	20.03	20	20.8	7.53	05/02/2022	17:29
10	N2663-02DUP	1	Solid	20.04	20	20.9	7.54	05/02/2022	17:30
11	CCV2	1	Water	NA	NA	20.3	12.03	05/02/2022	17:32

WORKLIST(Hardcopy Internal Chain)

283119948

WorkList Name : corrosivity n2641

WorkList ID : 159208

Department : Wet-Chemistry

Date : 05-02-2022 14:41:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/29/2022	Solid	N2641-03	Corrosivity	Cool 4 deg C	PSEG03	L11	TP-21	04/29/2022	9045D
04/29/2022	Solid	N2641-06	Corrosivity	Cool 4 deg C	PSEG03	L11	TP-24	04/29/2022	9045D
04/29/2022	Solid	N2644-03	Corrosivity	Cool 4 deg C	PSEG03	L11	OILY-RAGS-DEBRIS	04/29/2022	9045D
05/02/2022	Solid	N2663-02	Corrosivity	Cool 4 deg C	PSEG03	K22	TP-25	05/02/2022	9045D

Date/Time 05/02/22 16:10
 Raw Sample Received by: JD WCI
 Raw Sample Relinquished by: JD CSM

Date/Time 05/02/22 18:00
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JD WCI