



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119949

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	05/02/2022	17:10
2	CAL2	1	Water	NA	NA	20.4	7.00	05/02/2022	17:11
3	CAL3	1	Water	NA	NA	20.4	10.03	05/02/2022	17:12
4	ICV	1	Water	NA	NA	20.5	6.99	05/02/2022	17:15
5	CCV1	1	Water	NA	NA	20.3	2.02	05/02/2022	17:35
6	N2644-05	1	Water	NA	NA	20.3	7.30	05/02/2022	17:36
7	N2646-03	1	Water	NA	NA	20.1	5.99	05/02/2022	17:38
8	N2653-08	1	Water	NA	NA	20.4	7.32	05/02/2022	17:40
9	N2653-09	1	Water	NA	NA	20.6	6.95	05/02/2022	17:42
10	N2658-01	1	Water	NA	NA	20.7	8.66	05/02/2022	17:45
11	N2658-02	1	Water	NA	NA	20.5	8.71	05/02/2022	17:47
12	N2658-03	1	Water	NA	NA	20.2	8.28	05/02/2022	17:49
13	N2658-06	1	Water	NA	NA	20.4	5.56	05/02/2022	17:50
14	N2658-06DUP	1	Water	NA	NA	20.5	5.57	05/02/2022	17:51
15	CCV2	1	Water	NA	NA	20.4	12.03	05/02/2022	17:53

119949

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph w n2644

WorkList ID : 159211

Department : Wet-Chemistry

Date : 05-02-2022 15:56:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/29/2022	Water	N2644-05	pH	Cool 4 deg C	PSEG03	COOLANT		04/29/2022	9040C
04/29/2022	Water	N2646-03	pH	Cool 4 deg C	TETR06	L11	TT-042-IDWGW-20220429	04/29/2022	9040C
05/02/2022	Water	N2653-08	pH	Cool 4 deg C	PSEG03	K22	30992	05/02/2022	9040C
05/02/2022	Water	N2653-09	pH	Cool 4 deg C	PSEG03	K22	30991	05/02/2022	9040C
05/02/2022	Water	N2658-01	pH	Cool 4 deg C	PSEG03	K22	40817	05/02/2022	9040C
05/02/2022	Water	N2658-02	pH	Cool 4 deg C	PSEG03	K22	40818	05/02/2022	9040C
05/02/2022	Water	N2658-03	pH	Cool 4 deg C	PSEG03	K22	40819	05/02/2022	9040C
05/02/2022	Water	N2658-06	pH	Cool 4 deg C	PSEG03	K22	31751	05/02/2022	9040C

Date/Time 05/02/22 16:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 05/02/22 18:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]