


LB119960

7470A

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.4723e-004

C= 3.4991e-002

Rho= 0.9992217

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.041	0.041	26	0.000	26					
0.2	0.200	0.233	0.033	801	0.0 %	801					
2.5	2.500	2.472	-0.028	9859	0.0 %	9859					
5.0	5.000	5.066	0.066	20350	0.0 %	20350					
7.5	7.500	7.206	-0.294	29005	0.0 %	29005					
10.0	10.000	10.181	0.181	41038	0.0 %	41038					

0/00

17

-1

-4

2

LB119960 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 03 May 2022 12:06:38							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7470A	PPB	03 May 2022 12:15:14	S	Std
0.2 - 1	50.2	801	-	0.2000	7470A	PPB	03 May 2022 12:17:31	S	Std
2.5 - 1	52.5	9859	-	2.5000	7470A	PPB	03 May 2022 12:19:48	S	Std
5.0 - 1	55	20350	-	5.0000	7470A	PPB	03 May 2022 12:22:04	S	Std
7.5 - 1	57.5	29005	-	7.5000	7470A	PPB	03 May 2022 12:26:55	S	Std
10.0 - 1	610	41038	-	10.0000	7470A	PPB	03 May 2022 12:29:10	S	Std
ICV06 - 1	ICV06	17014	4.2414	-	7470A	PPB	03 May 2022 12:37:03	U	SMPL
ICB06 - 1	ICB06	-61	0.0199	-	7470A	PPB	03 May 2022 12:39:18	U	SMPL
CCV72 - 1	CCV72	20666	5.1443	-	7470A	PPB	03 May 2022 12:41:32	U	SMPL
CCB72 - 1	CCB72	-50	0.0226	-	7470A	PPB	03 May 2022 12:43:48	U	SMPL
CRA - 1	CRA	778	0.2273	-	7470A	PPB	03 May 2022 12:46:02	U	SMPL
HighStd - 1	HighStd	39124	9.7077	-	7470A	PPB	03 May 2022 12:48:17	U	SMPL
ChkStd - 1	ChkStd	27873	6.9261	-	7470A	PPB	03 May 2022 12:50:32	U	SMPL
PB144553BL - 1	PBW	-122	0.0048	-	7470A	PPB	03 May 2022 12:52:48	U	SMPL
PB144553BS - 1	LCSW	14004	3.4972	-	7470A	PPB	03 May 2022 12:55:03	U	SMPL
N2563-01 - 1	STORM-WATER-042622	77	0.0540	-	7470A	PPB	03 May 2022 12:57:20	U	SMPL
N2598-03 - 1	OW-TOTE-COMP	3885	0.9955	-	7470A	PPB	03 May 2022 13:04:38	U	SMPL
N2598-03DUP - 1	OW-TOTE-COMPDUP	3698	0.9493	-	7470A	PPB	03 May 2022 13:06:55	U	SMPL
N2598-03MS - 1	OW-TOTE-COMPMS	16003	3.9915	-	7470A	PPB	03 May 2022 13:09:12	U	SMPL
N2598-03MSD - 1	OW-TOTE-COMPMSD	14054	3.5096	-	7470A	PPB	03 May 2022 13:11:26	U	SMPL
CCV73 - 1	CCV73	20382	5.0741	-	7470A	PPB	03 May 2022 13:13:43	U	SMPL
CCB73 - 1	CCB73	-20	0.0300	-	7470A	PPB	03 May 2022 13:16:00	U	SMPL
N2619-24 - 1	SOIL-EQUIPMENT-RINSE	-20	0.0300	-	7470A	PPB	03 May 2022 13:18:15	U	SMPL
N2619-26 - 1	FIELD-BLANK	3	0.0357	-	7470A	PPB	03 May 2022 13:20:30	U	SMPL
N2624-01 - 1	N47992	315	0.1129	-	7470A	PPB	03 May 2022 13:22:45	U	SMPL
N2646-03 - 1	TT-042-IDWGW-20220429	18	0.0394	-	7470A	PPB	03 May 2022 13:25:01	U	SMPL
N2653-09 - 1	30991	178	0.0790	-	7470A	PPB	03 May 2022 13:27:17	U	SMPL
N2658-04 - 1	40817-19	208	0.0864	-	7470A	PPB	03 May 2022 13:29:33	U	SMPL
N2658-06 - 1	31751	270	0.1017	-	7470A	PPB	03 May 2022 13:31:49	U	SMPL
N2598-03LX5 - 1		592	0.1814	-	7470A	PPB	03 May 2022 13:34:05	U	SMPL
N2598-03A - 1		20328	5.0607	-	7470A	PPB	03 May 2022 13:36:22	U	SMPL
CCV74 - 1	CCV74	20650	5.1404	-	7470A	PPB	03 May 2022 13:38:38	U	SMPL
CCB74 - 1	CCB74	-31	0.0273	-	7470A	PPB	03 May 2022 13:40:53	U	SMPL