

LB 119962


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

42342



A= 0.0000e+000

B= 2.3917e-004 slope

C= -1.0691e-001 y-intercept

Rho= 0.9997802

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/D
0.00	0.000	-0.017	-0.017	377	0.000	377					=
0.05	0.050					185	0				7
0.20	0.200	0.213	0.013	1336	0.0 %	1336					-2
2.50	2.500	2.452	-0.048	10701	0.0 %	10701					3
5.00	5.000	5.143	0.143	21950	0.0 %	21950					1
7.50	7.500	7.389	-0.111	31340	0.0 %	31340					-1
10.0	10.000	10.020	0.020	42342	0.0 %	42342					0

LB119962
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	377	-		0 SFAM01.1	PPB	5/3/2022 14:53	Std
	0.2 S01	1336	-		0.2 SFAM01.1	PPB	5/3/2022 14:55	Std
	2.5 S02	10701	-		2.5 SFAM01.1	PPB	5/3/2022 14:58	Std
	5 S03	21950	-		5 SFAM01.1	PPB	5/3/2022 15:00	Std
	7.5 S04	31340	-		7.5 SFAM01.1	PPB	5/3/2022 15:02	Std
	10 S05	42342	-		10 SFAM01.1	PPB	5/3/2022 15:06	Std
ICV007	ICV007	16132	3.7514 -		SFAM01.1	PPB	5/3/2022 15:09	SMPL
ICB007	ICB007	337	-0.0263 -		SFAM01.1	PPB	5/3/2022 15:11	SMPL
CCV075	CCV075	20416	4.776 -		SFAM01.1	PPB	5/3/2022 15:13	SMPL
CCB075	CCB075	169	-0.0665 -		SFAM01.1	PPB	5/3/2022 15:15	SMPL
PB144560BL	PBW560	-34	-0.115 -		SFAM01.1	PPB	5/3/2022 15:18	SMPL
N2497-03	A0BJ1	-16	-0.1107 -		SFAM01.1	PPB	5/3/2022 15:22	SMPL
N2497-04	A0BJ3	23	-0.1014 -		SFAM01.1	PPB	5/3/2022 15:24	SMPL
N2497-05	A0BJ5	33	-0.099 -		SFAM01.1	PPB	5/3/2022 15:27	SMPL
N2497-06	A0BJ7	40	-0.0973 -		SFAM01.1	PPB	5/3/2022 15:29	SMPL
N2497-07	A0BJ7D	30	-0.0997 -		SFAM01.1	PPB	5/3/2022 15:31	SMPL
N2497-08	A0BJ7S	4103	0.8744 -		SFAM01.1	PPB	5/3/2022 15:36	SMPL
N2497-09	A0BL5	12	-0.104 -		SFAM01.1	PPB	5/3/2022 15:38	SMPL
N2497-10	A0BL7	30	-0.0997 -		SFAM01.1	PPB	5/3/2022 15:41	SMPL
N2497-11	A0BM3	26	-0.1007 -		SFAM01.1	PPB	5/3/2022 15:43	SMPL
N2497-12	A0BM6	11	-0.1043 -		SFAM01.1	PPB	5/3/2022 15:45	SMPL
N2497-13	A0BM8	14	-0.1036 -		SFAM01.1	PPB	5/3/2022 15:47	SMPL
N2497-14	A0BN0	1	-0.1067 -		SFAM01.1	PPB	5/3/2022 15:50	SMPL
N2497-15	A0BN2	-10	-0.1093 -		SFAM01.1	PPB	5/3/2022 15:52	SMPL
N2497-16	A0BN4	507	0.0143 -		SFAM01.1	PPB	5/3/2022 15:54	SMPL
N2497-17	A0BN8	1748	0.3112 -		SFAM01.1	PPB	5/3/2022 15:56	SMPL
N2497-18	A0BP0	6	-0.1055 -		SFAM01.1	PPB	5/3/2022 15:59	SMPL
N2497-19	A0BQ2	22	-0.1016 -		SFAM01.1	PPB	5/3/2022 16:01	SMPL
N2497-20	A0BQ4	384	-0.0151 -		SFAM01.1	PPB	5/3/2022 16:03	SMPL
CCV076	CCV076	21486	5.0319 -		SFAM01.1	PPB	5/3/2022 16:05	SMPL
CCB076	CCB076	336	-0.0265 -		SFAM01.1	PPB	5/3/2022 16:08	SMPL
PB144561BL	PBW561	239	-0.0497 -		SFAM01.1	PPB	5/3/2022 16:10	SMPL
N2503-03	A0BJ2	345	-0.0244 -		SFAM01.1	PPB	5/3/2022 16:14	SMPL
N2503-04	A0BJ4	415	-0.0077 -		SFAM01.1	PPB	5/3/2022 16:17	SMPL
N2503-05	A0BJ6	448	0.0002 -		SFAM01.1	PPB	5/3/2022 16:19	SMPL
N2503-06	A0BJ8	452	0.0012 -		SFAM01.1	PPB	5/3/2022 16:21	SMPL
N2503-07	A0BJ8D	426	-0.005 -		SFAM01.1	PPB	5/3/2022 16:23	SMPL
N2503-08	A0BJ8S	4632	1.0009 -		SFAM01.1	PPB	5/3/2022 16:26	SMPL
N2503-09	A0BL6	311	-0.0325 -		SFAM01.1	PPB	5/3/2022 16:28	SMPL
N2503-10	A0BL8	661	0.0512 -		SFAM01.1	PPB	5/3/2022 16:30	SMPL
N2503-11	A0BM7	712	0.0634 -		SFAM01.1	PPB	5/3/2022 16:33	SMPL
N2503-12	A0BM9	698	0.06 -		SFAM01.1	PPB	5/3/2022 16:35	SMPL
N2503-13	A0BN1	413	-0.0081 -		SFAM01.1	PPB	5/3/2022 16:37	SMPL
N2503-14	A0BN3	376	-0.017 -		SFAM01.1	PPB	5/3/2022 16:39	SMPL
N2503-15	A0BN5	149	-0.0713 -		SFAM01.1	PPB	5/3/2022 16:42	SMPL
N2503-16	A0BN9	486	0.0093 -		SFAM01.1	PPB	5/3/2022 16:44	SMPL

LB119962
INSTRUMENT ID : CV1

N2503-17	A0BP1	296	-0.0361 -	SFAM01.1	PPB	5/3/2022 16:46 SMPL
N2503-18	A0BP3	175	-0.0651 -	SFAM01.1	PPB	5/3/2022 16:48 SMPL
CCV077	CCV077	21613	5.0623 -	SFAM01.1	PPB	5/3/2022 16:51 SMPL
CCB077	CCB077	321	-0.0301 -	SFAM01.1	PPB	5/3/2022 16:53 SMPL
N2497-01	A0BE2	4394	0.944 -	SFAM01.1	PPB	5/3/2022 16:57 SMPL
N2503-01	A0BE4	4472	0.9627 -	SFAM01.1	PPB	5/3/2022 17:04 SMPL
CCV078	CCV078	21689	5.0805 -	SFAM01.1	PPB	5/3/2022 17:06 SMPL
CCB078	CCB078	328	-0.0285 -	SFAM01.1	PPB	5/3/2022 17:09 SMPL