



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB119963

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP96488
calibration std. phosphate 0.5 ppm	WP96489
calibration std. phosphate 0.3 ppm	WP96490
calibration std. phosphate 0.1 ppm	WP96491
calibration std. phosphate 0.05 ppm	WP96492
calibration std. 0 ppm	WP96493
phosphate CCV std.	WP96494
Combined reagent	WP96572
Phenolphthalein indicator	WP93767
Sodium hydroxide, 1N	WP95405
Phosphate ICV-LCS Std	WP96495

Intercept: -0.0018

Slope: 0.6518

Regression: 0.999933

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.003		05/03/2022	13:10
2	CAL2	0.05	1	50	50	0.031	0.05	0	05/03/2022	13:10
3	CAL3	0.10	1	50	50	0.065	0.102	-2	05/03/2022	13:11
4	CAL4	0.30	1	50	50	0.188	0.291	3	05/03/2022	13:11
5	CAL5	0.50	1	50	50	0.325	0.501	-0.2	05/03/2022	13:12
6	CAL6	1.00	1	50	50	0.651	1.002	-0.2	05/03/2022	13:12



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB119963

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.322	0.497	05/03/2022	13:13
2	ICB		1	50	50	0.001	0.004	05/03/2022	13:13
3	CCV1	0.50	1	50	50	0.321	0.495	05/03/2022	13:14
4	CCB1		1	50	50	0.003	0.007	05/03/2022	13:14
5	PB144556BL		1	50	50	0.002	0.006	05/03/2022	13:15
6	PB144556BS	0.50	1	50	50	0.324	0.500	05/03/2022	13:15
7	N2590-01		1	50	50	0.739	1.137	05/03/2022	13:16
8	N2590-01DUP		1	50	50	0.737	1.133	05/03/2022	13:16
9	N2590-01MS	0.50	1	50	50	0.819	1.259	05/03/2022	13:17
10	N2590-01MSD	0.50	1	50	50	0.818	1.258	05/03/2022	13:17
11	N2593-01		1	50	50	0.754	1.160	05/03/2022	13:18
12	N2636-01		1	50	50	1.397	2.146	05/03/2022	13:18
13	N2590-01		5	50	50	0.212	0.328	05/03/2022	13:19
14	N2590-01DUP		5	50	50	0.213	0.330	05/03/2022	13:19
15	CCV2	0.50	1	50	50	0.320	0.494	05/03/2022	13:20
16	CCB2		1	50	50	0.003	0.007	05/03/2022	13:20
17	N2590-01MS	0.50	5	50	50	0.270	0.417	05/03/2022	13:21
18	N2590-01MSD	0.50	5	50	50	0.271	0.419	05/03/2022	13:21
19	N2593-01		10	50	50	0.181	0.280	05/03/2022	13:22
20	N2636-01		10	50	50	0.408	0.629	05/03/2022	13:22
21	CCV3	0.50	1	50	50	0.324	0.500	05/03/2022	13:23
22	CCB3		1	50	50	0.001	0.004	05/03/2022	13:23