



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119965

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/03/2022	16:25
2	CAL2	1	Water	NA	NA	20.3	7.00	05/03/2022	16:26
3	CAL3	1	Water	NA	NA	20.3	10.03	05/03/2022	16:28
4	ICV	1	Water	NA	NA	20.4	6.99	05/03/2022	16:30
5	CCV1	1	Water	NA	NA	20.4	2.01	05/03/2022	16:32
6	N2674-01	1	Solid	20.03	20	21.6	6.84	05/03/2022	16:35
7	N2674-01DUP	1	Solid	20.04	20	21.7	6.85	05/03/2022	16:36
8	N2675-01	1	Solid	20.02	20	21.2	7.68	05/03/2022	16:39
9	CCV2	1	Water	NA	NA	20.4	12.03	05/03/2022	16:40

UB 119965

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-N2674 WorkList ID : 159250 Department : Wet-Chemistry Date : 05-03-2022 16:13:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/03/2022	Solid	N2674-01	pH <i>OK</i>	Cool 4 deg C	PSEG05	K22	BU-2-050322	05/03/2022	9045D
05/03/2022	Solid	N2675-01	pH <i>OK</i>	Cool 4 deg C	PSEG05	K22	BU-1-050322	05/03/2022	9045D

Date/Time 05/03/22 16:20
Raw Sample Received by: *JP(WC)*
Raw Sample Relinquished by: *JP(WC)*

Date/Time 05/03/22 14:30
Raw Sample Received by: *JP(WC)*
Raw Sample Relinquished by: *JP(WC)*