


LB119972

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.2400e-004 slope

C= 8.5644e-003 intercept

Rho= 0.9994950

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.005	0.005	-17	0.000	-17					=
0.05	0.050					185	0				-3
0.20	0.200	0.194	-0.006	829	0.0 %	829					-5
2.50	2.500	2.370	-0.130	10544	0.0 %	10544					5
5.00	5.000	5.243	0.243	23370	0.0 %	23370					1
7.50	7.500	7.439	-0.061	33170	0.0 %	33170					1
10.0	10.000	9.949	-0.051	44376	0.0 %	44376					

LB119972
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-17	-		0 SFAM01.1	PPB	5/4/2022 11:46	Std
	0.2 S01	829	-		0.2 SFAM01.1	PPB	5/4/2022 11:48	Std
	2.5 S02	10544	-		2.5 SFAM01.1	PPB	5/4/2022 11:51	Std
	5 S03	23370	-		5 SFAM01.1	PPB	5/4/2022 11:53	Std
	7.5 S04	33170	-		7.5 SFAM01.1	PPB	5/4/2022 11:55	Std
	10 S05	44376	-		10 SFAM01.1	PPB	5/4/2022 12:02	Std
ICV008	ICV008	17987	4.0376 -		SFAM01.1	PPB	5/4/2022 12:07	SMPL
ICB008	ICB008	-91	-0.0118 -		SFAM01.1	PPB	5/4/2022 12:09	SMPL
CCV079	CCV079	22448	5.0369 -		SFAM01.1	PPB	5/4/2022 12:12	SMPL
CCB079	CCB079	-152	-0.0255 -		SFAM01.1	PPB	5/4/2022 12:14	SMPL
PB144578BL	PBW678	-11	0.0061 -		SFAM01.1	PPB	5/4/2022 12:19	SMPL
N2504-01	A0BN6	2011	0.459 -		SFAM01.1	PPB	5/4/2022 12:21	SMPL
N2504-02	A0BN6D	1807	0.4133 -		SFAM01.1	PPB	5/4/2022 12:23	SMPL
N2504-03	A0BN6S	6087	1.372 -		SFAM01.1	PPB	5/4/2022 12:25	SMPL
N2504-04	A0BF5	1194	0.276 -		SFAM01.1	PPB	5/4/2022 12:28	SMPL
N2504-05	A0BG3	1849	0.4227 -		SFAM01.1	PPB	5/4/2022 12:30	SMPL
N2504-06	A0BG5	1734	0.397 -		SFAM01.1	PPB	5/4/2022 12:32	SMPL
N2504-07	A0BK3	1799	0.4115 -		SFAM01.1	PPB	5/4/2022 12:34	SMPL
N2504-08	A0BK7	1489	0.3421 -		SFAM01.1	PPB	5/4/2022 12:37	SMPL
N2504-09	A0BK9	1264	0.2917 -		SFAM01.1	PPB	5/4/2022 12:39	SMPL
N2504-10	A0BL1	664	0.1573 -		SFAM01.1	PPB	5/4/2022 12:41	SMPL
N2504-11	A0BL3	714	0.1685 -		SFAM01.1	PPB	5/4/2022 12:43	SMPL
N2504-12	A0BQ8	904	0.2111 -		SFAM01.1	PPB	5/4/2022 12:46	SMPL
N2504-13	A0BR0	722	0.1703 -		SFAM01.1	PPB	5/4/2022 12:48	SMPL
N2504-14	A0BE3	22403	5.0268 -		SFAM01.1	PPB	5/4/2022 12:50	SMPL
N2504-16	A0BF3	492	0.1188 -		SFAM01.1	PPB	5/4/2022 12:52	SMPL
N2504-17	A0BG7	580	0.1385 -		SFAM01.1	PPB	5/4/2022 12:55	SMPL
N2504-18	A0BG9	562	0.1345 -		SFAM01.1	PPB	5/4/2022 12:57	SMPL
N2504-19	A0BH1	575	0.1374 -		SFAM01.1	PPB	5/4/2022 12:59	SMPL
N2504-20	A0BH3	733	0.1728 -		SFAM01.1	PPB	5/4/2022 13:02	SMPL
CCV080	CCV080	22066	4.9513 -		SFAM01.1	PPB	5/4/2022 13:04	SMPL
CCB080	CCB080	-51	-0.0029 -		SFAM01.1	PPB	5/4/2022 13:06	SMPL
N2504-21	A0BH5	533	0.128 -		SFAM01.1	PPB	5/4/2022 13:08	SMPL
PB144579BL	PBW679	275	0.0702 -		SFAM01.1	PPB	5/4/2022 13:11	SMPL
N2505-01	A0BN7	271	0.0693 -		SFAM01.1	PPB	5/4/2022 13:13	SMPL
N2505-02	A0BN7D	290	0.0735 -		SFAM01.1	PPB	5/4/2022 13:15	SMPL
N2505-03	A0BN7S	4611	1.0414 -		SFAM01.1	PPB	5/4/2022 13:17	SMPL
N2505-04	A0BE1	22272	4.9974 -		SFAM01.1	PPB	5/4/2022 13:20	SMPL
N2505-06	A0BF4	226	0.0592 -		SFAM01.1	PPB	5/4/2022 13:22	SMPL
N2505-07	A0BG8	338	0.0843 -		SFAM01.1	PPB	5/4/2022 13:24	SMPL
N2505-08	A0BH0	299	0.0755 -		SFAM01.1	PPB	5/4/2022 13:26	SMPL
N2505-09	A0BH2	316	0.0793 -		SFAM01.1	PPB	5/4/2022 13:29	SMPL
N2505-10	A0BH4	287	0.0729 -		SFAM01.1	PPB	5/4/2022 13:31	SMPL
N2505-11	A0BK8	296	0.0749 -		SFAM01.1	PPB	5/4/2022 13:33	SMPL
N2505-12	A0BL0	284	0.0722 -		SFAM01.1	PPB	5/4/2022 13:35	SMPL
N2505-13	A0BL2	312	0.0785 -		SFAM01.1	PPB	5/4/2022 13:38	SMPL

LB119972
INSTRUMENT ID : CV1

N2505-14	A0BL4	336	0.0838 -	SFAM01.1	PPB	5/4/2022 13:40 SMPL
N2505-15	A0BQ9	344	0.0856 -	SFAM01.1	PPB	5/4/2022 13:42 SMPL
N2505-16	A0BR1	348	0.0865 -	SFAM01.1	PPB	5/4/2022 13:44 SMPL
N2505-18	A0BF6	358	0.0888 -	SFAM01.1	PPB	5/4/2022 13:47 SMPL
N2505-19	A0BG4	271	0.0693 -	SFAM01.1	PPB	5/4/2022 13:49 SMPL
N2505-20	A0BG6	255	0.0657 -	SFAM01.1	PPB	5/4/2022 13:51 SMPL
CCV081	CCV081	21929	4.9206 -	SFAM01.1	PPB	5/4/2022 13:54 SMPL
CCB081	CCB081	-52	-0.0031 -	SFAM01.1	PPB	5/4/2022 13:56 SMPL
N2505-21	A0BK4	298	0.0753 -	SFAM01.1	PPB	5/4/2022 13:58 SMPL
CCV082	CCV082	22192	4.9795 -	SFAM01.1	PPB	5/4/2022 14:12 SMPL
CCB082	CCB082	-76	-0.0085 -	SFAM01.1	PPB	5/4/2022 14:14 SMPL