



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119985

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/04/2022	16:20
2	CAL2	1	Water	NA	NA	20.7	7.00	05/04/2022	16:21
3	CAL3	1	Water	NA	NA	20.7	10.03	05/04/2022	16:22
4	ICV	1	Water	NA	NA	20.6	6.98	05/04/2022	16:25
5	CCV1	1	Water	NA	NA	20.6	2.01	05/04/2022	16:30
6	N2694-04	1	Water	NA	NA	20.4	8.43	05/04/2022	16:32
7	N2694-05	1	Water	NA	NA	20.4	5.43	05/04/2022	16:35
8	N2694-06	1	Water	NA	NA	20.2	7.54	05/04/2022	16:36
9	N2694-06DUP	1	Water	NA	NA	20.3	7.55	05/04/2022	16:37
10	CCV2	1	Water	NA	NA	20.7	12.03	05/04/2022	16:39

LB 119985

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n2694 WATER WorkList ID : 159274 Department : Wet-Chemistry Date : 05-04-2022 14:52:04

WORKLIST NAME : CONCERNY RESCUE 100											
---	--	--	--	--	--	--	--	--	--	--	--

Date/Time 05/04/22 16:57
Raw Sample Received by: JH (CSC)
Raw Sample Relinquished by: JH (CSC)

Date/Time 05/04/22 18:00
Raw Sample Received by: JH (CSC)
Raw Sample Relinquished by: JH (CSC)