



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119986

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/04/2022	16:20
2	CAL2	1	Water	NA	NA	20.7	7.00	05/04/2022	16:21
3	CAL3	1	Water	NA	NA	20.7	10.03	05/04/2022	16:22
4	ICV	1	Water	NA	NA	20.6	6.98	05/04/2022	16:25
5	CCV1	1	Water	NA	NA	20.6	2.01	05/04/2022	16:30
6	N2684-01	1	Solid	20.01	20	21.2	5.05	05/04/2022	16:32
7	N2684-03	1	Solid	20.02	20	21.3	7.38	05/04/2022	16:33
8	N2685-07	1	Solid	20.04	20	21.3	12.22	05/04/2022	16:38
9	N2694-01	1	Solid	20.01	20	21.3	6.48	05/04/2022	16:40
10	N2694-02	1	Solid	20.04	20	21.7	6.94	05/04/2022	16:42
11	N2694-03	1	Solid	20.03	20	21.4	5.05	05/04/2022	16:45
12	N2703-02	1	Solid	20.03	20	21.4	9.77	05/04/2022	16:47
13	N2703-05	1	Solid	20.04	20	21.0	8.80	05/04/2022	16:49
14	N2703-08	1	Solid	20.02	20	20.8	8.11	05/04/2022	17:00
15	N2703-08DUP	1	Solid	20.03	20	20.9	8.12	05/04/2022	17:01
16	CCV2	1	Water	NA	NA	20.6	12.03	05/04/2022	17:05

WB 119986

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n2694 S

WorkList ID : 159273

Department : Wet-Chemistry

Date : 05-04-2022 14:51:12

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
05/04/2022	Solid	N2684-01	Corrosivity	Cool 4 deg C	PSEG03	K23 BT-OIL	05/04/2022	9045D
05/04/2022	Solid	N2684-03	Corrosivity	Cool 4 deg C	PSEG03	K23 BT-OILY-DEB	05/04/2022	9045D
05/04/2022	Solid	N2685-07	Corrosivity	Cool 4 deg C	ROUX02	K23 WC-14B-5	05/04/2022	9045D
05/04/2022	Solid	N2694-01	Corrosivity	Cool 4 deg C	PSEG03	K23 1952-1943-COMP1	05/04/2022	9045D
05/04/2022	Solid	N2694-02	Corrosivity	Cool 4 deg C	PSEG03	K23 1944-1950-COMP2	05/04/2022	9045D
05/04/2022	Solid	N2694-03	Corrosivity	Cool 4 deg C	PSEG03	K23 1949	05/04/2022	9045D
05/04/2022	Solid	N2703-02	Corrosivity	Cool 4 deg C	PSEG03	K23 TP-28	05/04/2022	9045D
05/04/2022	Solid	N2703-05	Corrosivity	Cool 4 deg C	PSEG03	K23 MH-29S	05/04/2022	9045D
05/04/2022	Solid	N2703-08	Corrosivity	Cool 4 deg C	PSEG03	K23 MH-29D	05/04/2022	9045D

Date/Time 05/04/22 16:10
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)

Date/Time 05/04/22 18:20
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)