

LB119990


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42453



A= 0.0000e+000

B= 2.3293e-004

C= 4.7112e-002

Rho= 0.9989972

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.038	0.038	-37	0.000	-37					13
0.2	0.200	0.226	0.026	768	0.0 %	768					2
2.5	2.500	2.557	0.057	10777	0.0 %	10777					-6
5.0	5.000	4.695	-0.305	19954	0.0 %	19954					3
7.5	7.500	7.747	0.247	33058	0.0 %	33058					-1
10.0	10.000	9.936	-0.064	42453	0.0 %	42453					

LB119990 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 May 2022 10:02:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-37	-	0.0000	7470A	PPB	05 May 2022 10:41:49	S	Std
0.2 - 1	50.2	768	-	0.2000	7470A	PPB	05 May 2022 10:44:06	S	Std
2.5 - 1	52.5	10777	-	2.5000	7470A	PPB	05 May 2022 10:46:23	S	Std
5.0 - 1	55	19954	-	5.0000	7470A	PPB	05 May 2022 10:48:40	S	Std
7.5 - 1	57.5	33058	-	7.5000	7470A	PPB	05 May 2022 10:50:56	S	Std
10.0 - 1	510	42453	-	10.0000	7470A	PPB	05 May 2022 10:53:12	S	Std
ICV10 - 1	ICV10	15758	3.7176	-	7470A	PPB	05 May 2022 10:58:07	U	SMPL
ICB10 - 1	ICB10	-51	0.0352	-	7470A	PPB	05 May 2022 11:00:22	U	SMPL
CCV87 - 1	CCV87	20894	4.9140	-	7470A	PPB	05 May 2022 11:02:41	U	SMPL
CCB87 - 1	CCB87	-80	0.0285	-	7470A	PPB	05 May 2022 11:04:56	U	SMPL
CRA - 1	CRA	815	0.2370	-	7470A	PPB	05 May 2022 11:07:11	U	SMPL
HighStd - 1	HighStd	42321	9.9050	-	7470A	PPB	05 May 2022 11:09:26	U	SMPL
ChkStd - 1	ChkStd	32038	7.5098	-	7470A	PPB	05 May 2022 11:11:41	U	SMPL
PB144595BL - 1	PBW	-131	0.0166	-	7470A	PPB	05 May 2022 11:13:58	U	SMPL
PB144595BS - 1	LCSW	16077	3.7919	-	7470A	PPB	05 May 2022 11:16:15	U	SMPL
N1443-05 - 1	TP-1-SOUTH-COMP	-81	0.0282	-	7470A	PPB	05 May 2022 11:18:31	U	SMPL
N2641-03 - 1	TP-21	-15	0.0436	-	7470A	PPB	05 May 2022 11:20:47	U	SMPL
N2641-06 - 1	TP-24	31	0.0543	-	7470A	PPB	05 May 2022 11:23:04	U	SMPL
N2644-03 - 1	OILY-RAGS-DEBRIS	-13	0.0441	-	7470A	PPB	05 May 2022 11:34:32	U	SMPL
N2663-02 - 1	TP-25	41	0.0567	-	7470A	PPB	05 May 2022 11:36:46	U	SMPL
CCV88 - 1	CCV88	20437	4.8075	-	7470A	PPB	05 May 2022 11:39:02	U	SMPL
CCB88 - 1	CCB88	-66	0.0317	-	7470A	PPB	05 May 2022 11:41:17	U	SMPL
N2663-02DUP - 1	TP-25DUP	4	0.0480	-	7470A	PPB	05 May 2022 11:43:33	U	SMPL
N2663-02MS - 1	TP-25MS	15621	3.6857	-	7470A	PPB	05 May 2022 11:45:48	U	SMPL
N2663-02MSD - 1	TP-25MSD	15797	3.7267	-	7470A	PPB	05 May 2022 11:48:03	U	SMPL
PB144524TB - 1	PB144524TB	-60	0.0331	-	7470A	PPB	05 May 2022 11:50:19	U	SMPL
N2663-02LX5 - 1		22	0.0522	-	7470A	PPB	05 May 2022 11:52:36	U	SMPL
N2663-02A - 1		15272	3.6044	-	7470A	PPB	05 May 2022 11:54:53	U	SMPL
CCV89 - 1	CCV89	20598	4.8450	-	7470A	PPB	05 May 2022 11:57:09	U	SMPL
CCB89 - 1	CCB89	-31	0.0399	-	7470A	PPB	05 May 2022 11:59:25	U	SMPL