

LB 118999

=====

Test results	Aquakem 7.2AQ1	Page: 1
--------------	----------------	---------

=====

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

5/5/2022 8:41

Reviewed by : <u>12</u>	Instrument ID : Konelab
-------------------------	-------------------------

Test: Total CN

Sample Id	Result <i>up/c</i>	Dil. 1 +	Response	Errors
ICV1	96.025	0.0	0.080	
ICB1	0.477	0.0	0.001	
CCV1	248.867	0.0	0.206	
CCB1	0.669	0.0	0.001	
PB144548BL	0.606	0.0	0.001	
PB144548BS	96.878	0.0	0.081	
N2656-05	1495.299	0.0	1.236	Test limit high
N2656-05DLX5	305.185	0.0	0.253	
CCV2	253.854	0.0	0.210	
CCB2	0.866	0.0	0.001	

N	10
Mean	249.873
SD	453.3429
CV%	181.43

Aquakem v. 7.2AQ1

Results from time period:

Wed May 04 17:04:49 2022

Wed May 04 17:11:36 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.2993	µg/l	5/4/2022 14:14:35	
5.0PPBCN	A	Total CN	P	5.1508	µg/l	5/4/2022 14:14:36	
10PPBCN	A	Total CN	P	10.104	µg/l	5/4/2022 14:14:37	
50PPBCN	A	Total CN	P	48.6771	µg/l	5/4/2022 14:14:38	
100PPBCN	A	Total CN	P	98.9935	µg/l	5/4/2022 14:14:39	
250PPBCN	A	Total CN	P	252.8906	µg/l	5/4/2022 14:14:40	
500PPBCN	A	Total CN	P	498.8847	µg/l	5/4/2022 14:14:41	
ICV1	S	Total CN	P	96.0247	µg/l	5/4/2022 17:04:50	
ICB1	S	Total CN	P	0.4772	µg/l	5/4/2022 17:04:51	
CCV1	S	Total CN	P	248.8674	µg/l	5/4/2022 17:04:54	
CCB1	S	Total CN	P	0.6692	µg/l	5/4/2022 17:04:56	
PB144548BL	S	Total CN	P	0.6061	µg/l	5/4/2022 17:04:57	
PB144548BS	S	Total CN	P	96.878	µg/l	5/4/2022 17:04:59	
N2656-05	S	Total CN	P	1495.299	µg/l	5/4/2022 17:11:30	
N2656-05DLX5	S	Total CN	P	305.1851	µg/l	5/4/2022 17:11:32	
CCV2	S	Total CN	P	253.8542	µg/l	5/4/2022 17:11:34	
CCB2	S	Total CN	P	0.8662	µg/l	5/4/2022 17:11:35	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

5/4/2022 14:15


Test Total CN

Accepted 5/4/2022 14:15

Factor 1211
Bias 0.001

Coeff. of det. 0.999939

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.2993	0.0000	-
2	5.0PPBCN	0.005	5.1508	5.0000	3.0
3	10PPBCN	0.009	10.1040	10.0000	1.0
4	50PPBCN	0.041	48.6771	50.0000	-2.6
5	100PPBCN	0.082	98.9935	100.0000	-1.0
6	250PPBCN	0.209	252.8906	250.0000	1.2
7	500PPBCN	0.413	498.8847	500.0000	-0.2

12 5/4/22