

LB119995


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46617



A= 0.0000e+000

B= 2.1549e-004 slope

C= 2.7290e-002 y-intercept

Rho= 0.9998785

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.038	0.038	52	0.000	52					=
0.05	0.050					185	0				12
0.20	0.200	0.224	0.024	913	0.0 %	913					0
2.50	2.500	2.502	0.002	11482	0.0 %	11482					-2
5.00	5.000	4.892	-0.108	22575	0.0 %	22575					0
7.50	7.500	7.471	-0.029	34545	0.0 %	34545					0
10.0	10.000	10.073	0.073	46617	0.0 %	46617					1

LB119995
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	52	-		0 SFAM01.1	PPB	5/5/2022 12:37	Std
	0.2 S01	913	-		0.2 SFAM01.1	PPB	5/5/2022 12:40	Std
	2.5 S02	11482	-		2.5 SFAM01.1	PPB	5/5/2022 12:42	Std
	5 S03	22575	-		5 SFAM01.1	PPB	5/5/2022 12:44	Std
	7.5 S04	34545	-		7.5 SFAM01.1	PPB	5/5/2022 12:47	Std
	10 S05	46617	-		10 SFAM01.1	PPB	5/5/2022 12:49	Std
ICV011	ICV011	17750	3.8522	-	SFAM01.1	PPB	5/5/2022 12:52	SMPL
ICB011	ICB011	-37	0.0193	-	SFAM01.1	PPB	5/5/2022 12:54	SMPL
CCV090	CCV90	22480	4.8715	-	SFAM01.1	PPB	5/5/2022 12:56	SMPL
CCB090	CCB090	-23	0.0223	-	SFAM01.1	PPB	5/5/2022 12:58	SMPL
PB144598BL	PBS598	38	0.0355	-	SFAM01.1	PPB	5/5/2022 13:01	SMPL
N2700-01	BGBP8	1730	0.4001	-	SFAM01.1	PPB	5/5/2022 13:03	SMPL
N2700-02	BGBP8D	1773	0.4094	-	SFAM01.1	PPB	5/5/2022 13:05	SMPL
N2700-03	BGBP8S	12522	2.7256	-	SFAM01.1	PPB	5/5/2022 13:10	SMPL
N2700-04	BGBP9	1492	0.3488	-	SFAM01.1	PPB	5/5/2022 13:12	SMPL
N2700-05	BGBQ0	1660	0.385	-	SFAM01.1	PPB	5/5/2022 13:14	SMPL
N2700-06	BGBQ1	1579	0.3675	-	SFAM01.1	PPB	5/5/2022 13:17	SMPL
N2700-07	BGBQ2	1449	0.3395	-	SFAM01.1	PPB	5/5/2022 13:19	SMPL
CCV091	CCV091	23070	4.9986	-	SFAM01.1	PPB	5/5/2022 13:21	SMPL
CCB091	CCB091	-38	0.0191	-	SFAM01.1	PPB	5/5/2022 13:23	SMPL