

Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB120001

BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP96488
calibration std. phosphate 0.5 ppm	WP96489
calibration std. phosphate 0.3 ppm	WP96490
calibration std. phosphate 0.1 ppm	WP96491
calibration std. phosphate 0.05 ppm	WP96492
calibration std. 0 ppm	WP96493
phosphate CCV std.	WP96494
Combined reagent	WP96617
Sodium hydroxide, 1N	WP95405
Phosphate ICV-LCS Std	WP96495

Intercept: -0.0012

Slope: 0.6523

Regression: 0.999931

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.002		05/05/2022	13:15
2	CAL2	0.05	1	50	50	0.033	0.052	-4	05/05/2022	13:15
3	CAL3	0.10	1	50	50	0.066	0.103	-3	05/05/2022	13:16
4	CAL4	0.30	1	50	50	0.189	0.292	2.7	05/05/2022	13:16
5	CAL5	0.50	1	50	50	0.324	0.499	0.2	05/05/2022	13:17
6	CAL6	1.00	1	50	50	0.653	1.003	-0.3	05/05/2022	13:17



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB120001

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.499	05/05/2022	13:18
2	ICB		1	50	50	0.001	0.003	05/05/2022	13:18
3	CCV1	0.50	1	50	50	0.328	0.505	05/05/2022	13:19
4	CCB1		1	50	50	0.000	0.002	05/05/2022	13:19
5	PB144612BL		1	1.00	50	0.002	0.005	05/05/2022	13:20
6	PB144612BS	0.50	1	1.00	50	0.321	0.494	05/05/2022	13:20
7	N2656-08		1	1.01	50	2.154	3.304	05/05/2022	13:21
8	N2656-08		50	1.01	50	0.502	0.771	05/05/2022	13:21
9	CCV2	0.50	1	50	50	0.322	0.495	05/05/2022	13:22
10	CCB2		1	50	50	0.001	0.003	05/05/2022	13:22