



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 5/6/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:45
In Date: 05/05/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 05/06/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB120004

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2708-01	EXL29	1	0.93	8.62	9.55	8.49	87.7	
N2708-02	EXL30	2	0.92	8.79	9.71	9.2	94.2	
N2708-03	EXL30MS	3	0.92	8.79	9.71	9.2	94.2	
N2708-04	EXL30MSD	4	0.92	8.79	9.71	9.2	94.2	
N2708-05	EXL31	5	0.93	8.95	9.88	9.46	95.3	
N2708-06	EXL32	6	0.92	8.78	9.7	9.42	96.8	
N2708-07	EXL33	7	0.9	8.95	9.85	8.77	87.9	
N2708-08	EXL34	8	0.93	8.71	9.64	8.55	87.5	
N2708-09	EXL35	9	0.92	8.86	9.78	8.98	91.0	
N2708-10	EXL36	10	0.93	8.77	9.7	9.42	96.8	
N2708-11	EXL37	11	0.92	8.87	9.79	9.4	95.6	
N2708-12	EXL38	12	0.93	8.77	9.7	8.35	84.6	
N2708-13	EXL39	13	0.94	8.92	9.86	9.27	93.4	
N2708-14	EXL40	14	0.93	8.91	9.84	9.6	97.3	
N2708-15	EXL41	15	0.93	8.87	9.8	9.46	96.2	
N2708-16	EXL42	16	0.92	8.78	9.7	8.27	83.7	
N2708-17	EXL43	17	0.93	8.75	9.68	9.27	95.3	
N2708-18	EXL44	18	0.92	8.89	9.81	9.49	96.4	
N2708-19	EXL45	19	0.93	8.87	9.8	9.6	97.7	
N2708-20	EXL47	20	0.92	8.82	9.74	8.85	89.9	
N2708-21	EXL48	21	0.91	8.99	9.9	9.51	95.7	
N2708-22	EXL49	22	0.92	8.98	9.9	9.63	97.0	
N2708-23	VHBLK001	23	1.00	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$