


LB 120006

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.7827e-004

C= -1.4979e-002

Rho= 0.9999313

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.010	-0.010	28	0.000	28					-
0.2	0.200	0.189	-0.011	1144	0.0 %	1144					-6
2.5	2.500	2.500	0.000	14110	0.0 %	14110					0
5.0	5.000	5.078	0.078	28567	0.0 %	28567					2
7.5	7.500	7.431	-0.069	41765	0.0 %	41765					-1
10.0	10.000	10.012	0.012	56245	0.0 %	56245					0

LB120006 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 May 2022 14:48:07

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		28	-	0.0000	7471B	PPB	05 May 2022 14:50:54	S	Std
0.2 - 1		1144	-	0.2000	7471B	PPB	05 May 2022 14:53:12	S	Std
2.5 - 1		14110	-	2.5000	7471B	PPB	05 May 2022 14:55:28	S	Std
5.0 - 1		28567	-	5.0000	7471B	PPB	05 May 2022 14:57:45	S	Std
7.5 - 1		41765	-	7.5000	7471B	PPB	05 May 2022 15:00:01	S	Std
10.0 - 1		56245	-	10.0000	7471B	PPB	05 May 2022 15:02:17	S	Std
ICV12 - 1	ICV12	21474	3.8133	-7471B	PPB		05 May 2022 15:05:01	U	SMPL
ICB12 - 1	ICB12	-61	-0.0259	-7471B	PPB		05 May 2022 15:07:16	U	SMPL
CCV92 - 1	CCV92	28580	5.0801	-7471B	PPB		05 May 2022 15:12:22	U	SMPL
CCB92 - 1	CCB92	-86	-0.0303	-7471B	PPB		05 May 2022 15:14:36	U	SMPL
CRA - 1	CRA	1119	0.1845	-7471B	PPB		05 May 2022 15:16:52	U	SMPL
HighStd - 1	HighStd	55901	9.9508	-7471B	PPB		05 May 2022 15:19:07	U	SMPL
ChkStd - 1	ChkStd	39508	7.0283	-7471B	PPB		05 May 2022 15:21:22	U	SMPL
PB144582BL - 1	PBS	-106	-0.0339	-7471B	PPB		05 May 2022 15:23:39	U	SMPL
PB144582BS - 1	LCSS	19719	3.5004	-7471B	PPB		05 May 2022 15:25:55	U	SMPL
N2665-01 - 1	SB-12	445	0.0644	-7471B	PPB		05 May 2022 15:28:11	U	SMPL
N2665-01DUP - 1	SB-12	363	0.0497	-7471B	PPB		05 May 2022 15:30:28	U	SMPL
N2665-01MS - 1	SB-12	20801	3.6933	-7471B	PPB		05 May 2022 15:32:44	U	SMPL
N2665-01MSD - 1	SB-12	20701	3.6755	-7471B	PPB		05 May 2022 15:35:01	U	SMPL
N2665-02 - 1	SB-12	362	0.0496	-7471B	PPB		05 May 2022 15:37:16	U	SMPL
CCV93 - 1	CCV93	26917	4.7836	-7471B	PPB		05 May 2022 15:39:31	U	SMPL
CCB93 - 1	CCB93	-55	-0.0248	-7471B	PPB		05 May 2022 15:41:47	U	SMPL
N2665-03 - 1	SB-8	962	0.1565	-7471B	PPB		05 May 2022 15:44:02	U	SMPL
N2665-04 - 1	SB-8	256	0.0307	-7471B	PPB		05 May 2022 15:46:17	U	SMPL
N2665-05 - 1	SB-4	259	0.0312	-7471B	PPB		05 May 2022 15:48:32	U	SMPL
N2665-06 - 1	SB-4	173	0.0159	-7471B	PPB		05 May 2022 15:50:47	U	SMPL
N2665-07 - 1	SB-4	424	0.0606	-7471B	PPB		05 May 2022 15:53:03	U	SMPL
N2665-08 - 1	SB-11	321	0.0422	-7471B	PPB		05 May 2022 15:55:19	U	SMPL
N2665-09 - 1	SB-11	404	0.0570	-7471B	PPB		05 May 2022 15:57:36	U	SMPL
N2665-10 - 1	SB-10	279	0.0348	-7471B	PPB		05 May 2022 15:59:52	U	SMPL
N2665-11 - 1	SB-10	196	0.0200	-7471B	PPB		05 May 2022 16:02:09	U	SMPL
N2665-12 - 1	SB-9	662	0.1030	-7471B	PPB		05 May 2022 16:04:24	U	SMPL
CCV94 - 1	CCV94	26943	4.7883	-7471B	PPB		05 May 2022 16:06:39	U	SMPL
CCB94 - 1	CCB94	-39	-0.0219	-7471B	PPB		05 May 2022 16:08:55	U	SMPL
N2665-13 - 1	SB-9	319	0.0419	-7471B	PPB		05 May 2022 16:11:10	U	SMPL
PB144583BL - 1	PBS	5	-0.0141	-7471B	PPB		05 May 2022 16:13:26	U	SMPL
PB144583BS - 1	LCSS	20290	3.6022	-7471B	PPB		05 May 2022 16:15:41	U	SMPL
N2677-01 - 1	SB-5	514	0.0767	-7471B	PPB		05 May 2022 16:17:56	U	SMPL
N2677-01DUP - 1	SB-5DUP	417	0.0594	-7471B	PPB		05 May 2022 16:20:12	U	SMPL
N2677-01MS - 1	SB-5MS	21436	3.8065	-7471B	PPB		05 May 2022 16:22:28	U	SMPL
N2677-01MSD - 1	SB-5MSD	20649	3.6662	-7471B	PPB		05 May 2022 16:24:43	U	SMPL
N2677-02 - 1	SB-5	452	0.0656	-7471B	PPB		05 May 2022 16:27:00	U	SMPL
N2677-03 - 1	SB-1	195	0.0198	-7471B	PPB		05 May 2022 16:29:18	U	SMPL
N2677-04 - 1	SB-1	110	0.0046	-7471B	PPB		05 May 2022 16:31:33	U	SMPL
CCV95 - 1	CCV95	28910	5.1389	-7471B	PPB		05 May 2022 16:33:48	U	SMPL
CCB95 - 1	CCB95	107	0.0041	-7471B	PPB		05 May 2022 16:36:04	U	SMPL
N2677-05 - 1	SB-2	379	0.0526	-7471B	PPB		05 May 2022 16:38:19	U	SMPL
N2677-06 - 1	SB-2	338	0.0453	-7471B	PPB		05 May 2022 16:40:34	U	SMPL
N2677-07 - 1	SB-3	230	0.0260	-7471B	PPB		05 May 2022 16:42:49	U	SMPL
N2677-08 - 1	SB-3	278	0.0346	-7471B	PPB		05 May 2022 16:45:05	U	SMPL
N2677-09 - 1	SB-6	386	0.0538	-7471B	PPB		05 May 2022 16:47:20	U	SMPL
N2677-10 - 1	SB-6	120	0.0064	-7471B	PPB		05 May 2022 16:49:37	U	SMPL
N2677-11 - 1	SB-7	479	0.0704	-7471B	PPB		05 May 2022 16:51:54	U	SMPL
N2677-12 - 1	SB-7	238	0.0275	-7471B	PPB		05 May 2022 16:54:10	U	SMPL
PB144589BL - 1	PBS	-2	-0.0153	-7471B	PPB		05 May 2022 16:56:28	U	SMPL
PB144589BS - 1	LCSS	22423	3.9825	-7471B	PPB		05 May 2022 16:58:45	U	SMPL
CCV96 - 1	CCV96	28949	5.1459	-7471B	PPB		05 May 2022 17:01:00	U	SMPL
CCB96 - 1	CCB96	125	0.0073	-7471B	PPB		05 May 2022 17:03:18	U	SMPL
N2666-01 - 1	WASTE	3963	0.6915	-7471B	PPB		05 May 2022 17:05:34	U	SMPL
N2680-01 - 1	B1-0-2	13336	2.3625	-7471B	PPB		05 May 2022 17:07:49	U	SMPL
N2680-02 - 1	B1-5-7	468	0.0685	-7471B	PPB		05 May 2022 17:10:05	U	SMPL
N2680-03 - 1	B2-0-2	77795	13.8539	-7471B	PPB		05 May 2022 17:12:21	U	SMPL
N2680-04 - 1	B2-5-7	3598	0.6265	-7471B	PPB		05 May 2022 17:14:37	U	SMPL
N2680-05 - 1	B3-0-2	43783	7.7904	-7471B	PPB		05 May 2022 17:16:54	U	SMPL
N2680-06 - 1	B3-6-8	291	0.0369	-7471B	PPB		05 May 2022 17:19:12	U	SMPL
N2680-07 - 1	B4-0-2	50295	8.9513	-7471B	PPB		05 May 2022 17:21:31	U	SMPL
N2680-08 - 1	B4-5-7	1580	0.2667	-7471B	PPB		05 May 2022 17:23:50	U	SMPL
N2680-09 - 1	B5-0-2	11251	1.9908	-7471B	PPB		05 May 2022 17:26:07	U	SMPL
CCV97 - 1	CCV97	28746	5.1097	-7471B	PPB		05 May 2022 17:28:23	U	SMPL
CCB97 - 1	CCB97	-6	-0.0160	-7471B	PPB		05 May 2022 17:30:39	U	SMPL
N2677-01 - 1	B5-5-7	418	0.0595	-7471B	PPB		05 May 2022 17:32:57	U	SMPL
N2680-10DUP - 1	B5-5-7DUP	388	0.0542	-7471B	PPB		05 May 2022 17:35:13	U	SMPL
N2680-10MS - 1	B5-5-7MS	21606	3.8368	-7471B	PPB		05 May 2022 17:37:29	U	SMPL

LB120006 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 May 2022 14:48:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
N2680-10MSD - 1	B5-5-7MSD	22327	3.9654		-7471B	PPB	05 May 2022 17:39:45	U	SMPL
N2689-01 - 1	OK-01-050422	1553	0.2619		-7471B	PPB	05 May 2022 17:42:01	U	SMPL
N2691-01 - 1	DEEP-WC-1	47505	8.4540		-7471B	PPB	05 May 2022 17:44:17	U	SMPL
N2691-03 - 1	SHALLOW-WC-1	22958	4.0779		-7471B	PPB	05 May 2022 17:46:34	U	SMPL
N2691-05 - 1	SHALLOW-WC-2	18176	3.2253		-7471B	PPB	05 May 2022 17:48:52	U	SMPL
N2692-01 - 1	72-12002	5076	0.8899		-7471B	PPB	05 May 2022 17:51:10	U	SMPL
CCV98 - 1	CCV98	26878	4.7767		-7471B	PPB	05 May 2022 17:53:26	U	SMPL
CCB98 - 1	CCB98	-125	-0.0373		-7471B	PPB	05 May 2022 17:55:42	U	SMPL
N2665-01LX5 - 1		-62	-0.0260		-7471B	PPB	05 May 2022 17:57:58	U	SMPL
N2665-01A - 1		22805	4.0506		-7471B	PPB	05 May 2022 18:00:16	U	SMPL
N2677-01LX5 - 1		14	-0.0125		-7471B	PPB	05 May 2022 18:02:33	U	SMPL
N2677-01A - 1		23209	4.1226		-7471B	PPB	05 May 2022 18:04:49	U	SMPL
N2680-10LX5 - 1		4	-0.0143		-7471B	PPB	05 May 2022 18:07:04	U	SMPL
N2680-10A - 1		21883	3.8862		-7471B	PPB	05 May 2022 18:09:23	U	SMPL
CCV99 - 1	CCV99	27540	4.8947		-7471B	PPB	05 May 2022 18:11:39	U	SMPL
CCB99 - 1	CCB99	-25	-0.0194		-7471B	PPB	05 May 2022 18:13:57	U	SMPL