



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB120014

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/05/2022	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	05/05/2022	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	05/05/2022	16:18
4	ICV	1	Water	NA	NA	20.4	7.00	05/05/2022	16:20
5	CCV1	1	Water	NA	NA	20.4	2.01	05/05/2022	16:22
6	N2656-02	1	Solid	20.02	20	21.1	4.57	05/05/2022	16:25
7	N2723-02	1	Solid	20.03	20	23.4	8.85	05/05/2022	16:30
8	N2723-04	1	Solid	20.04	20	23.1	8.83	05/05/2022	16:33
9	N2723-06	1	Solid	20.02	20	22.9	8.11	05/05/2022	16:35
10	N2723-06DUP	1	Solid	20.01	20	22.8	8.12	05/05/2022	16:36
11	CCV2	1	Water	NA	NA	20.4	12.03	05/05/2022	16:38

WORKLIST(Hardcopy Internal Chain)

120014

WorkList Name : corrosivity n2703

WorkList ID : 159323

Department : Wet-Chemistry

Date : 05-05-2022 13:55:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/05/2022	Solid	N2723-02	Corrosivity	Cool 4 deg C	PSEG03	K23	MMC-PIT1	05/05/2022	9045D
05/05/2022	Solid	N2723-04	Corrosivity	Cool 4 deg C	PSEG03	K23	MMC-PIT2	05/05/2022	9045D
05/05/2022	Solid	N2723-06	Corrosivity	Cool 4 deg C	PSEG03	K23	MMC-PIT3	05/05/2022	9045D

Date/Time 05/05/22 11:10
 Raw Sample Received by: JP(WC)
 Raw Sample Relinquished by: JP(WC)

Date/Time 05/05/22 18:20
 Raw Sample Received by: JP(WC)
 Raw Sample Relinquished by: JP(WC)