



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB120028

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	05/06/2022	08:25
2	CAL2	1	Water	NA	NA	20.4	7.01	05/06/2022	08:26
3	CAL3	1	Water	NA	NA	20.4	10.03	05/06/2022	08:27
4	ICV	1	Water	NA	NA	20.5	6.98	05/06/2022	08:30
5	CCV1	1	Water	NA	NA	20.5	2.01	05/06/2022	08:32
6	N2690-01	1	Water	NA	NA	20.6	7.39	05/06/2022	08:35
7	N2705-01	1	Water	NA	NA	20.3	6.83	05/06/2022	08:39
8	N2712-13	1	Water	NA	NA	22.3	6.66	05/06/2022	08:40
9	N2718-01	1	Water	NA	NA	20.1	7.23	05/06/2022	08:42
10	N2718-03	1	Water	NA	NA	20.8	6.96	05/06/2022	08:45
11	N2718-05	1	Water	NA	NA	21.1	7.18	05/06/2022	08:47
12	N2720-01	1	Water	NA	NA	22.0	4.55	05/06/2022	08:50
13	N2725-03	1	Water	NA	NA	22.1	5.67	05/06/2022	09:00
14	N2725-04	1	Water	NA	NA	22.1	5.88	05/06/2022	09:02
15	N2725-05	1	Water	NA	NA	22.3	5.84	05/06/2022	09:05
16	CCV2	1	Water	NA	NA	20.5	12.02	05/06/2022	09:10
17	N2725-06	1	Water	NA	NA	22.3	5.43	05/06/2022	09:15
18	N2725-07	1	Water	NA	NA	22.3	5.86	05/06/2022	09:20
19	N2725-08	1	Water	NA	NA	22.4	5.56	05/06/2022	09:25
20	N2725-09	1	Water	NA	NA	22.2	5.84	05/06/2022	09:26

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	N2726-01	1	Water	NA	NA	22.3	5.68	05/06/2022	09:30
22	N2726-01DUP	1	Water	NA	NA	22.4	5.69	05/06/2022	09:31
23	CCV3	1	Water	NA	NA	20.4	2.01	05/06/2022	09:35

18 120028

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2725 WorkList ID : 159341 Department : Wet-Chemistry Date : 05-06-2022 08:14:41

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
05/04/2022	Water	N2690-01	pH	Cool 4 deg C	PSEG04	K23 402	05/04/2022	9040C
05/05/2022	Water	N2705-01	pH	Cool 4 deg C	PSEG03	N48967	05/05/2022	9040C
05/05/2022	Water	N2712-13	pH	Cool 4 deg C	PSEG03	K23 GROUND-WATER	05/05/2022	9040C
05/05/2022	Water	N2718-01	pH	Cool 4 deg C	PSEG04	K23 DSN002	05/05/2022	9040C
05/05/2022	Water	N2718-03	pH	Cool 4 deg C	PSEG04	K23 DSN001	05/05/2022	9040C
05/05/2022	Water	N2718-05	pH	Cool 4 deg C	PSEG04	K23 RWS003	05/05/2022	9040C
05/03/2022	Water	N2720-01	pH	Cool 4 deg C	STER04	K23 POOL-WATER	05/03/2022	9040C
05/05/2022	Water	N2725-03	pH C	Cool 4 deg C	PSEG03	K23 TRE-1276	05/05/2022	9040C
05/05/2022	Water	N2725-04	pH C	Cool 4 deg C	PSEG03	K23 TRE-1277	05/05/2022	9040C
05/05/2022	Water	N2725-05	pH C	Cool 4 deg C	PSEG03	K23 TRE-1278	05/05/2022	9040C
05/05/2022	Water	N2725-06	pH C	Cool 4 deg C	PSEG03	K23 TRE-1279	05/05/2022	9040C
05/05/2022	Water	N2725-07	pH C	Cool 4 deg C	PSEG03	K23 TRE-1280	05/05/2022	9040C
05/05/2022	Water	N2725-08	pH C	Cool 4 deg C	PSEG03	K23 TRE-1281	05/05/2022	9040C
05/05/2022	Water	N2725-09	pH C	Cool 4 deg C	PSEG03	K23 TRE-1593	05/05/2022	9040C
05/05/2022	Water	N2726-01	pH C	Cool 4 deg C	PSEG03	K23 FLOOR-DRAINING-TOTE	05/05/2022	9040C

Date/Time 05/06/22 08:20
Raw Sample Received by: JDCM
Raw Sample Relinquished by: JDCMDate/Time 05/06/22 12:00
Raw Sample Received by: JDCM
Raw Sample Relinquished by: JDCM