



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB120029

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	05/06/2022	08:25
2	CAL2	1	Water	NA	NA	20.4	7.01	05/06/2022	08:26
3	CAL3	1	Water	NA	NA	20.4	10.03	05/06/2022	08:27
4	ICV	1	Water	NA	NA	20.5	6.98	05/06/2022	08:30
5	CCV1	1	Water	NA	NA	20.4	2.02	05/06/2022	09:42
6	N2725-01	1	Solid	20.02	20	21.9	7.81	05/06/2022	09:45
7	N2725-02	1	Solid	20.03	20	21.6	6.44	05/06/2022	09:48
8	N2725-02DUP	1	Solid	20.02	20	21.7	6.45	05/06/2022	09:49
9	CCV2	1	Water	NA	NA	20.2	12.02	05/06/2022	09:50

VB 120029

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2725 s WorkList ID : 159354 Department : Wet-Chemistry Date : 05-06-2022 09:23:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Customer Sample	Collect Date	Method
						Storage	Location			
05/05/2022	Solid	N2725-01	pH	Cool 4 deg C	PSEG03	K23	TRE-1254		05/05/2022	9045D
05/05/2022	Solid	N2725-02	pH	Cool 4 deg C	PSEG03	K23	TRE-1284		05/05/2022	9045D

Date/Time 05/06/22 08:25
Raw Sample Received by: TP (WC)
Raw Sample Relinquished by: JD (CSM)

Date/Time 05/06/22 12:00
Raw Sample Received by: JD (CSM)
Raw Sample Relinquished by: TP (WC)