

LB120092


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44984



A= 0.0000e+000

B= 2.2169e-004 *sbpr*C= 5.6192e-002 *y=10128cept*

Rho= 0.9998602

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.053	0.053	-15	0.000	-15					11
0.05	0.050					185	0				18
0.20	0.200	0.236	0.036	812	0.0 %	812					15
2.50	2.500	2.371	-0.129	10443	0.0 %	10443					10
5.00	5.000	5.031	0.031	22440	0.0 %	22440					0
7.50	7.500	7.480	-0.020	33489	0.0 %	33489					0
10.0	10.000	10.029	0.029	44984	0.0 %	44984					0

LB120032
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-15	-		0 SFAM01.1	PPB	5/6/2022 14:35	Std
	0.2 S01	812	-	0.2	SFAM01.1	PPB	5/6/2022 14:38	Std
	2.5 S02	10443	-	2.5	SFAM01.1	PPB	5/6/2022 14:40	Std
	5 S03	22440	-	5	SFAM01.1	PPB	5/6/2022 14:44	Std
	7.5 S04	33489	-	7.5	SFAM01.1	PPB	5/6/2022 14:47	Std
	10 S05	44984	-	10	SFAM01.1	PPB	5/6/2022 14:49	Std
ICV013	ICV013	16685	3.755 -		SFAM01.1	PPB	5/6/2022 14:52	SMPL
ICB013	ICB013	-43	0.0467 -		SFAM01.1	PPB	5/6/2022 14:57	SMPL
CCV001	CCV001	22344	5.0096 -		SFAM01.1	PPB	5/6/2022 15:00	SMPL
CCB001	CCB001	85	0.075 -		SFAM01.1	PPB	5/6/2022 15:02	SMPL
PB144637BL	PBW637	134	0.0859 -		SFAM01.1	PPB	5/6/2022 15:05	SMPL
N2539-01	A0BH7	138	0.0868 -		SFAM01.1	PPB	5/6/2022 15:08	SMPL
N2539-02	A0BH9	162	0.0921 -		SFAM01.1	PPB	5/6/2022 15:10	SMPL
N2539-03	A0BJ9	202	0.101 -		SFAM01.1	PPB	5/6/2022 15:13	SMPL
N2539-04	A0BJ9D	160	0.0917 -		SFAM01.1	PPB	5/6/2022 15:15	SMPL
N2539-05	A0BJ9S	4482	1.0498 -		SFAM01.1	PPB	5/6/2022 15:18	SMPL
N2539-06	A0BK1	382	0.1409 -		SFAM01.1	PPB	5/6/2022 15:23	SMPL
N2539-07	A0BM0	-267	-0.003 -		SFAM01.1	PPB	5/6/2022 15:26	SMPL
N2539-08	A0BM1	362	0.1364 -		SFAM01.1	PPB	5/6/2022 15:29	SMPL
N2539-09	A0BP8	97	0.0777 -		SFAM01.1	PPB	5/6/2022 15:31	SMPL
N2539-10	A0BQ0	278	0.1178 -		SFAM01.1	PPB	5/6/2022 15:35	SMPL
PB144638BL	PBW638	17	0.06 -		SFAM01.1	PPB	5/6/2022 15:37	SMPL
N2540-01	A0BH6	79	0.0737 -		SFAM01.1	PPB	5/6/2022 15:39	SMPL
N2540-02	A0BH8	119	0.0826 -		SFAM01.1	PPB	5/6/2022 15:42	SMPL
N2540-03	A0BJ0	324	0.128 -		SFAM01.1	PPB	5/6/2022 15:44	SMPL
N2540-04	A0BK0	-51	0.0449 -		SFAM01.1	PPB	5/6/2022 15:47	SMPL
N2540-05	A0BK0D	-14	0.0531 -		SFAM01.1	PPB	5/6/2022 15:49	SMPL
N2540-06	A0BK0S	4307	1.011 -		SFAM01.1	PPB	5/6/2022 15:52	SMPL
N2540-07	A0BK2	-19	0.052 -		SFAM01.1	PPB	5/6/2022 15:54	SMPL
N2540-08	A0BP9	149	0.0892 -		SFAM01.1	PPB	5/6/2022 15:57	SMPL
CCV002	CCV002	22477	5.039 -		SFAM01.1	PPB	5/6/2022 15:59	SMPL
CCB002	CCB002	43	0.0657 -		SFAM01.1	PPB	5/6/2022 16:01	SMPL
N2540-09	A0BQ1	509	0.169 -		SFAM01.1	PPB	5/6/2022 16:04	SMPL
CCV003	CCV003	21887	4.9083 -		SFAM01.1	PPB	5/6/2022 16:07	SMPL
CCB003	CCB003	130	0.085 -		SFAM01.1	PPB	5/6/2022 16:09	SMPL