



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB120038

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/06/2022	09:25
2	N2653-01		1	4	21.00	69.8	69.8	05/06/2022	09:35
3	N2653-01DUP		1	4	22.00	71.6	71.6	05/06/2022	09:50
4	N2653-02		1	4	19.00	66.2	66.2	05/06/2022	10:02
5	N2653-03		1	3	20.00	68.0	68.0	05/06/2022	10:15
6	N2653-04		1	5	28.00	82.4	82.4	05/06/2022	10:35
7	N2653-05		1	4	20.00	68.0	68.0	05/06/2022	10:45
8	N2653-06		1	5	20.00	68.0	68.0	05/06/2022	10:55
9	N2653-07		1	4	19.00	66.2	66.2	05/06/2022	11:10
10	N2656-06		1	7	52.00	125.6	125.6	05/06/2022	11:35
11	N2684-01		1	10	100.00	>212.0	>212.0	05/06/2022	12:05
12	N2694-03		1	12	100.00	>212.0	>212.0	05/06/2022	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP5222-1

WorkList ID : 159216

Department : Wet-Chemistry

Date : 05-02-2022 16:55:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/09/2022	Solid	N2653-01	Flash Point	Cool 4 deg C	PSEG03	K22	31591	05/02/2022	1010B
05/09/2022	Solid	N2653-02	Flash Point	Cool 4 deg C	PSEG03	K22	31592	05/02/2022	1010B
05/09/2022	Solid	N2653-03	Flash Point	Cool 4 deg C	PSEG03	K22	31593	05/02/2022	1010B
05/09/2022	Solid	N2653-04	Flash Point	Cool 4 deg C	PSEG03	K22	31594	05/02/2022	1010B
05/09/2022	Solid	N2653-05	Flash Point	Cool 4 deg C	PSEG03	K22	31595	05/02/2022	1010B
05/09/2022	Solid	N2653-06	Flash Point	Cool 4 deg C	PSEG03	K22	31596	05/02/2022	1010B
05/09/2022	Solid	N2653-07	Flash Point	Cool 4 deg C	PSEG03	K22	31597	05/02/2022	1010B
05/09/2022	Solid	N2684-01	Flash Point	Cool 4 deg C	PSEG03	K23	BT-OIL	05/04/2022	1010B
05/11/2022	Solid	N2694-03	Flash Point	Cool 4 deg C	PSEG03	K23	1949	05/04/2022	1010B

Date/Time 05/06/22 09:05
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 05/06/22 12:45
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: KS (WC)