



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : ketankumar

Run Number: LB120049

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/06/2022	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	05/06/2022	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	05/06/2022	16:17
4	ICV	1	Water	NA	NA	20.4	6.99	05/06/2022	16:20
5	CCV1	1	Water	NA	NA	20.4	2.01	05/06/2022	16:35
6	N2730-01	1	Water	NA	NA	21.5	5.66	05/06/2022	16:36
7	N2734-01	1	Water	NA	NA	20.6	6.99	05/06/2022	16:37
8	N2734-02	1	Water	NA	NA	20.9	7.16	05/06/2022	16:39
9	N2734-05	1	Water	NA	NA	21.3	7.42	05/06/2022	16:40
10	N2738-05	1	Water	NA	NA	21.6	7.83	05/06/2022	16:43
11	N2741-01	1	Water	NA	NA	21.0	6.23	05/06/2022	16:45
12	N2741-01DUP	1	Water	NA	NA	21.1	6.24	05/06/2022	16:46
13	CCV2	1	Water	NA	NA	20.5	12.02	05/06/2022	16:47

120049

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2741 WorkList ID : 159377 Department : Wet-Chemistry Date : 05-06-2022 14:54:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/05/2022	Water	N2730-01	pH	Cool 4 deg C	STER14	L11	MINI-CELL	05/05/2022	9040C
05/06/2022	Water	N2734-01	pH	Cool 4 deg C	PSEG03	K23	RL-TOTE-A	05/06/2022	9040C
05/06/2022	Water	N2734-02	pH	Cool 4 deg C	PSEG03	K23	RL-TOTE-B	05/06/2022	9040C
05/06/2022	Water	N2734-05	pH	Cool 4 deg C	PSEG03	K23	FMI-120	05/06/2022	9040C
05/06/2022	Water	N2738-05	pH	Cool 4 deg C	PSEG03	K23	2466	05/06/2022	9040C
05/06/2022	Water	N2741-01	pH	Cool 4 deg C	loui01	K23	TWP01	05/06/2022	9040C

Date/Time 05/08/22 16:10
 Raw Sample Received by: JPC
 Raw Sample Relinquished by: JPC

Date/Time 05/06/22 18:20
 Raw Sample Received by: JPC
 Raw Sample Relinquished by: JPC