

LB120055


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

40512



A= 0.0000e+000

B= 2.4863e-004

C= -4.4743e-002

Rho= 0.9993671

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.013	-0.013	126	0.000	126					
0.2	0.200	0.181	-0.019	907	0.0 %	907					-10
2.5	2.500	2.704	0.204	11055	0.0 %	11055					8
5.0	5.000	4.758	-0.242	19316	0.0 %	19316					-5
7.5	7.500	7.543	0.043	30518	0.0 %	30518					1
10.0	10.000	10.028	0.028	40512	0.0 %	40512					0

LB120055 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 May 2022 08:02:02

Sample ID	Extended ID	μ Abs.	Conc.	Std. Cond	Method	Units	Date	Type	Type
0.0 - 1	50	126	-	0.0007470A	PPB		09 May 2022 08:30:37	S	Std
0.2 - 1	50.2	907	-	0.20007470A	PPB		09 May 2022 08:32:54	S	Std
2.5 - 1	50.5	11055	-	2.50007470A	PPB		09 May 2022 08:35:11	S	Std
5.0 - 1	50.5	19316	-	5.00007470A	PPB		09 May 2022 08:37:28	S	Std
7.5 - 1	50.5	30518	-	7.50007470A	PPB		09 May 2022 08:39:44	S	Std
10.0 - 1	510	40512	-	10.00007470A	PPB		09 May 2022 08:47:15	S	Std
ICV14 - 1	ICV14	16142	3.9687	-7470A	PPB		09 May 2022 08:52:29	U	SMPL
ICB14 - 1	ICB14	-19	-0.0495	-7470A	PPB		09 May 2022 08:54:44	U	SMPL
CCV04 - 1	CCV04	19368	4.7708	-7470A	PPB		09 May 2022 08:56:59	U	SMPL
CCB04 - 1	CCB04	36	-0.0358	-7470A	PPB		09 May 2022 08:59:14	U	SMPL
CRA - 1	CRA	905	0.1803	-7470A	PPB		09 May 2022 09:01:29	U	SMPL
HighStd - 1	HighStd	41009	10.1514	-7470A	PPB		09 May 2022 09:03:44	U	SMPL
ChkStd - 1	ChkStd	26613	6.5721	-7470A	PPB		09 May 2022 09:05:59	U	SMPL
PB144648BL - 1	PBW	52	-0.0318	-7470A	PPB		09 May 2022 09:08:15	U	SMPL
PB144648BS - 1	LCSW	15852	3.8966	-7470A	PPB		09 May 2022 09:10:32	U	SMPL
N2684-01 - 1	BT-OIL	276	0.0239	-7470A	PPB		09 May 2022 09:12:48	U	SMPL
N2684-03 - 1	BT-OILY-DEB	60	-0.0298	-7470A	PPB		09 May 2022 09:15:05	U	SMPL
N2685-07 - 1	WC-14B-5	154	-0.0065	-7470A	PPB		09 May 2022 09:17:21	U	SMPL
N2694-01 - 1	1952-1943-COMP1	246	0.0164	-7470A	PPB		09 May 2022 09:19:38	U	SMPL
N2694-02 - 1	1944-1950-COMP2	176	-0.0010	-7470A	PPB		09 May 2022 09:21:53	U	SMPL
CCV05 - 1	CCV05	19697	4.8526	-7470A	PPB		09 May 2022 09:24:08	U	SMPL
CCB05 - 1	CCB05	53	-0.0316	-7470A	PPB		09 May 2022 09:26:24	U	SMPL
N2694-03 - 1	1949	100	-0.0199	-7470A	PPB		09 May 2022 09:28:39	U	SMPL
N2703-02 - 1	TP-28	148	-0.0079	-7470A	PPB		09 May 2022 09:30:54	U	SMPL
N2703-05 - 1	MH-29S	115	-0.0162	-7470A	PPB		09 May 2022 09:33:10	U	SMPL
N2703-08 - 1	MH-29D	113	-0.0166	-7470A	PPB		09 May 2022 09:35:25	U	SMPL
N2706-01 - 1	RBS-BD002-01	161	-0.0047	-7470A	PPB		09 May 2022 09:37:41	U	SMPL
N2723-02 - 1	MMC-PIT1	169	-0.0027	-7470A	PPB		09 May 2022 09:39:57	U	SMPL
N2723-04 - 1	MMC-PIT2	197	0.0042	-7470A	PPB		09 May 2022 09:42:14	U	SMPL
N2723-06 - 1	MMC-PIT3	241	0.0152	-7470A	PPB		09 May 2022 09:44:31	U	SMPL
N2723-06DUP - 1	MMC-PIT3DUP	242	0.0154	-7470A	PPB		09 May 2022 09:46:48	U	SMPL
N2723-06MS - 1	MMC-PIT3MS	14498	3.5599	-7470A	PPB		09 May 2022 09:49:03	U	SMPL
CCV06 - 1	CCV06	19515	4.8073	-7470A	PPB		09 May 2022 09:51:18	U	SMPL
CCB06 - 1	CCB06	56	-0.0308	-7470A	PPB		09 May 2022 09:53:33	U	SMPL
N2723-06MSD - 1	MMC-PIT3MSD	15003	3.6855	-7470A	PPB		09 May 2022 09:58:38	U	SMPL
PB144670BL - 1	PBW	105	-0.0186	-7470A	PPB		09 May 2022 10:00:53	U	SMPL
PB144670BS - 1	LCSW	16745	4.1186	-7470A	PPB		09 May 2022 10:03:09	U	SMPL
N2694-04 - 1	1942	43	-0.0341	-7470A	PPB		09 May 2022 10:05:24	U	SMPL
N2694-05 - 1	1947-1945-1946-TOTE-COMP	28	-0.0378	-7470A	PPB		09 May 2022 10:07:39	U	SMPL
N2694-06 - 1	1951	-344	-0.1303	-7470A	PPB		09 May 2022 10:09:55	U	SMPL
N2694-06DUP - 1	1951DUP	147	-0.0082	-7470A	PPB		09 May 2022 10:12:11	U	SMPL
N2694-06MS - 1	1951MS	14280	3.5057	-7470A	PPB		09 May 2022 10:14:28	U	SMPL
N2694-06MSD - 1	1951MSD	14291	3.5085	-7470A	PPB		09 May 2022 10:16:45	U	SMPL
PB144650BL - 1	PBW	-11	-0.0475	-7470A	PPB		09 May 2022 10:19:01	U	SMPL
CCV07 - 1	CCV07	19224	4.7350	-7470A	PPB		09 May 2022 10:21:16	U	SMPL
CCB07 - 1	CCB07	36	-0.0358	-7470A	PPB		09 May 2022 10:23:31	U	SMPL
PB144650BS - 1	LCSW	16531	4.0654	-7470A	PPB		09 May 2022 10:25:47	U	SMPL
N2680-11 - 1	B1-GW	767	0.1460	-7470A	PPB		09 May 2022 10:32:53	U	SMPL
N2680-11DUP - 1	B1-GWDUP	711	0.1320	-7470A	PPB		09 May 2022 10:35:08	U	SMPL
N2680-11MS - 1	B1-GWMS	16199	3.9829	-7470A	PPB		09 May 2022 10:37:24	U	SMPL
N2680-11MSD - 1	B1-GWMSD	14645	3.5965	-7470A	PPB		09 May 2022 10:39:40	U	SMPL
N2680-12 - 1	B1-GW	-13	-0.0480	-7470A	PPB		09 May 2022 10:41:55	U	SMPL
N2680-13 - 1	B3-GW	340	0.0398	-7470A	PPB		09 May 2022 10:44:12	U	SMPL
N2680-14 - 1	B3-GW	42	-0.0343	-7470A	PPB		09 May 2022 10:46:29	U	SMPL
N2680-15 - 1	B4-GW	1695	0.3767	-7470A	PPB		09 May 2022 10:48:46	U	SMPL
N2680-16 - 1	B4-GW	17	-0.0405	-7470A	PPB		09 May 2022 10:51:02	U	SMPL
CCV08 - 1	CCV08	18398	4.5296	-7470A	PPB		09 May 2022 10:53:17	U	SMPL
CCB08 - 1	CCB08	85	-0.0236	-7470A	PPB		09 May 2022 10:55:32	U	SMPL
N2683-02 - 1	COMP	292	0.0279	-7470A	PPB		09 May 2022 10:57:48	U	SMPL
N2705-01 - 1	N48967	52	-0.0318	-7470A	PPB		09 May 2022 11:00:03	U	SMPL
N2712-13 - 1	GROUND-WATER	854	0.1676	-7470A	PPB		09 May 2022 11:02:19	U	SMPL
N2718-01 - 1	DSN002	373	0.0480	-7470A	PPB		09 May 2022 11:04:34	U	SMPL
N2718-03 - 1	DSN001	1037	0.2131	-7470A	PPB		09 May 2022 11:06:50	U	SMPL
N2718-05 - 1	RWS003	359	0.0445	-7470A	PPB		09 May 2022 11:09:06	U	SMPL
CCV09 - 1	CCV09	18707	4.6064	-7470A	PPB		09 May 2022 11:11:23	U	SMPL
CCB09 - 1	CCB09	69	-0.0276	-7470A	PPB		09 May 2022 11:13:40	U	SMPL
PB144609TB - 1	PB144609TB	154	-0.0065	-7470A	PPB		09 May 2022 11:15:57	U	SMPL
PB144611TB - 1	PB144611TB	183	0.0008	-7470A	PPB		09 May 2022 11:18:13	U	SMPL
N2723-06LX5 - 1		27	-0.0380	-7470A	PPB		09 May 2022 11:20:30	U	SMPL
N2723-06A - 1		16331	4.0157	-7470A	PPB		09 May 2022 11:22:46	U	SMPL
N2694-06LX5 - 1		-13	-0.0480	-7470A	PPB		09 May 2022 11:25:01	U	SMPL
N2694-06A - 1		16405	4.0341	-7470A	PPB		09 May 2022 11:27:17	U	SMPL
N2680-11LX5 - 1		179	-0.0002	-7470A	PPB		09 May 2022 11:29:33	U	SMPL
N2680-11A - 1		16558	4.0721	-7470A	PPB		09 May 2022 11:31:49	U	SMPL
CCV10 - 1	CCV10	19147	4.7158	-7470A	PPB		09 May 2022 11:36:50	U	SMPL
CCB10 - 1	CCB10	60	-0.0298	-7470A	PPB		09 May 2022 11:39:05	U	SMPL