



PERCENT SOLID

Supervisor: ketankumar

Analyst: jignesh

Date: 5/10/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:50

In Date: 05/09/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:20

Out Date: 05/10/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB120067

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2748-01	72-11940	6	1.00	1.00	2.00	2.00	100.0	plastic sample
N2748-02	72-11991	7	1.00	1.00	2.00	2.00	100.0	plastic sample
N2749-01	TP-1	8	0.9	8.61	9.51	8.54	88.7	
N2749-03	TP-1	9	0.9	8.61	9.51	8.54	88.7	
N2749-04	TP-2	10	0.92	8.74	9.66	8.61	88.0	
N2749-06	TP-2	11	0.92	8.74	9.66	8.61	88.0	
N2749-07	TP-3	12	0.91	8.82	9.73	8.46	85.6	
N2749-09	TP-3	13	0.91	8.82	9.73	8.46	85.6	
N2749-10	TP-4	14	0.91	8.74	9.65	8.55	87.4	
N2749-12	TP-4	15	0.91	8.74	9.65	8.55	87.4	
N2749-13	TP-5	16	0.95	9.01	9.96	8.74	86.5	
N2749-15	TP-5	17	0.95	9.01	9.96	8.74	86.5	
N2749-16	TP-6	18	0.92	8.71	9.63	8.37	85.5	
N2749-18	TP-6	19	0.92	8.70	9.62	8.37	85.6	
N2749-19	TP-7	20	0.91	8.82	9.73	8.81	89.6	
N2749-21	TP-7	21	0.91	8.82	9.73	8.81	89.6	
N2750-01	SV-RO-1/2	1	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N2750-02	SV-RO-1/1	2	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N2750-03	T-BU-12/8	3	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N2755-01	VNJ-210	22	0.9	8.90	9.8	9.29	94.3	
N2755-03	VNJ-210-EPH-2	23	0.9	8.90	9.8	9.29	94.3	
N2755-04	VNJ-221	24	0.94	8.61	9.55	9.00	93.6	
N2755-06	VNJ-221-EPH-2	25	0.94	8.61	9.55	9.00	93.6	
N2758-01	WASTE	26	0.95	9.02	9.97	8.87	87.8	
N2758-02	VOC	27	0.94	8.61	9.55	8.53	88.2	
N2758-03	1	28	0.92	9.02	9.94	8.92	88.7	
N2758-04	2	29	0.93	9.04	9.97	9.04	89.7	
N2758-05	3	30	0.94	9.00	9.94	9.27	92.6	



PERCENT SOLID

Supervisor: ketankumar
Analyst: jignesh
Date: 5/10/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:50
In Date: 05/09/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 05/10/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB120067

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2758-06	4	31	0.91	9.01	9.92	8.71	86.6	
N2758-07	5	32	0.94	9.02	9.96	9.02	89.6	
N2759-01	PL-02-050922	4	0.92	9.03	9.95	9.06	90.1	
N2759-02	PL-02-050922-EPH-2	5	0.92	9.03	9.95	9.06	90.1	
N2760-01	TP-34	33	0.94	8.93	9.87	9.13	91.7	
N2760-03	TP-34-EPH-2	34	0.94	9.02	9.96	9.18	91.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$