

LB 120068

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092Reviewed by : 12 Instrument ID : Konelab

5/9/2022 15:26

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.000	0.0	0.110	
ICB1	-0.002	0.0	0.006	
CCV1	0.984	0.0	0.108	
CCB1	-0.002	0.0	0.006	
PB144698BL	-0.003	0.0	0.006	
PB144698BS	0.989	0.0	0.109	
N2718-02	0.006	0.0	0.007	
N2718-02DUP	0.005	0.0	0.007	
N2718-02MS	1.034	0.0	0.114	
N2718-02MSD	1.021	0.0	0.112	
N2718-04	0.002	0.0	0.007	
N2718-06	0.010	0.0	0.008	
N2741-01	0.098	0.0	0.017	
CCV2	0.977	0.0	0.108	
CCB2	0.008	0.0	0.007	

N	15
Mean	0.408
SD	0.5015
CV%	122.76

Aquakem v. 7.2AQ1

Results from time period:

Mon May 09 14:42:48 2022

Mon May 09 15:26:07 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0035	mg/l	5/9/2022 14:42:48	
0.05PPM	A	Phenolics-	P	0.0512	mg/l	5/9/2022 14:42:49	
0.1PPM	A	Phenolics-	P	0.1006	mg/l	5/9/2022 14:42:50	
0.25PPM	A	Phenolics-	P	0.2522	mg/l	5/9/2022 14:42:51	
0.50PPM	A	Phenolics-	P	0.5055	mg/l	5/9/2022 14:42:52	
1.0PPM	A	Phenolics-	P	0.9772	mg/l	5/9/2022 14:42:53	
2.0PPM	A	Phenolics-	P	2.0097	mg/l	5/9/2022 14:42:54	
ICV1	S	Phenolics-	P	1.0003	mg/l	5/9/2022 15:18:46	
ICB1	S	Phenolics-	P	-0.0017	mg/l	5/9/2022 15:18:48	
CCV1	S	Phenolics-	P	0.9839	mg/l	5/9/2022 15:18:50	
CCB1	S	Phenolics-	P	-0.0021	mg/l	5/9/2022 15:18:51	
PB144698BL	S	Phenolics-	P	-0.0029	mg/l	5/9/2022 15:18:54	
PB144698BS	S	Phenolics-	P	0.9885	mg/l	5/9/2022 15:18:56	
N2718-02	S	Phenolics-	P	0.0062	mg/l	5/9/2022 15:25:59	
N2718-02DUP	S	Phenolics-	P	0.005	mg/l	5/9/2022 15:26:00	
N2718-02MS	S	Phenolics-	P	1.0343	mg/l	5/9/2022 15:26:01	
N2718-02MSD	S	Phenolics-	P	1.0215	mg/l	5/9/2022 15:26:02	
N2718-04	S	Phenolics-	P	0.0019	mg/l	5/9/2022 15:26:03	
N2718-06	S	Phenolics-	P	0.0102	mg/l	5/9/2022 15:26:04	
N2741-01	S	Phenolics-	P	0.0978	mg/l	5/9/2022 15:26:05	
CCV2	S	Phenolics-	P	0.9771	mg/l	5/9/2022 15:26:06	
CCB2	S	Phenolics-	P	0.0075	mg/l	5/9/2022 15:26:07	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

5/9/2022 14:43


Test Phenolics-

Accepted 5/9/2022 14:43

Factor 9.653
Bias 0.007

Coeff. of det. 0.999790

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.007	0.0035	0.0000	-
2	Phenol-2	0.012	0.0512	0.0500	2.4
3	Phenol-2	0.017	0.1006	0.1000	0.6
4	Phenol-2	0.033	0.2522	0.2500	0.9
5	Phenol-2	0.059	0.5055	0.5000	1.1
6	Phenol-2	0.108	0.9772	1.0000	-2.3
7	Phenol-2	0.215	2.0097	2.0000	0.5

12
5/9/22